

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012022\
 Data File : BF126617.D
 Acq On : 20 Jan 2022 18:06
 Operator : CG/JU
 Sample : N1189-11
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MTR-SOIL-6

Quant Time: Jan 21 04:14:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 19 07:04:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.963	152	110360	20.000	ng	0.00
21) Naphthalene-d8	8.245	136	426242	20.000	ng	0.00
39) Acenaphthene-d10	10.004	164	236894	20.000	ng	0.00
64) Phenanthrene-d10	11.498	188	429926	20.000	ng	0.00
76) Chrysene-d12	14.145	240	307861	20.000	ng	# 0.00
86) Perylene-d12	15.674	264	218665	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.581	112	718837	85.712	ng	0.00
7) Phenol-d6	6.575	99	1000326	85.848	ng	0.00
23) Nitrobenzene-d5	7.522	82	747468	68.639	ng	-0.01
42) 2,4,6-Tribromophenol	10.792	330	189579	85.513	ng	0.00
45) 2-Fluorobiphenyl	9.322	172	957752	61.918	ng	0.00
79) Terphenyl-d14	13.086	244	1104563	60.178	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012022\
Data File : BF126617.D
Acq On : 20 Jan 2022 18:06
Operator : CG/JU
Sample : N1189-11
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MTR-SOIL-6

Quant Time: Jan 21 04:14:33 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 19 07:04:14 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.963 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF126617.D
Acq: 20 Jan 2022 18:06




Instrument :
BNA_F
ClientSampleId :
MTR-SOIL-6


Tgt Ion:152 Resp: 110360
Ion Ratio Lower Upper
152 100
150 153.0 128.1 192.1
115 69.1 56.7 85.1



#5
2-Fluorophenol
Concen: 85.712 ng
RT: 5.581 min Scan# 594
Delta R.T. -0.000 min
Lab File: BF126617.D
Acq: 20 Jan 2022 18:06

Tgt Ion:112 Resp: 718837
Ion Ratio Lower Upper
112 100
64 70.1 49.5 74.3
63 34.9 23.9 35.9









#7
Phenol-d6
Concen: 85.848 ng
RT: 6.575 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF126617.D
Acq: 20 Jan 2022 18:06




Instrument :
BNA_F
ClientSampleId :
MTR-SOIL-6


Tgt Ion: 99 Resp: 1000326
Ion Ratio Lower Upper
99 100
42 16.6 14.2 21.4
71 41.5 24.3 36.5#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.245 min Scan# 1047
Delta R.T. -0.006 min
Lab File: BF126617.D
Acq: 20 Jan 2022 18:06

Tgt Ion:136 Resp: 426242
Ion Ratio Lower Upper
136 100
137 10.3 8.6 13.0
54 9.1 7.8 11.6
68 8.6 6.3 9.5









#23
Nitrobenzene-d5
Concen: 68.639 ng
RT: 7.522 min Scan# 91
Delta R.T. -0.012 min
Lab File: BF126617.D
Acq: 20 Jan 2022 18:06




Instrument :
BNA_F
ClientSampleId :
MTR-SOIL-6


Tgt Ion: 82 Resp: 747468
Ion Ratio Lower Upper
82 100
128 29.6 32.7 49.1#
54 41.6 41.7 62.5#



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BF126617.D
Acq: 20 Jan 2022 18:06

Tgt Ion: 164 Resp: 236894
Ion Ratio Lower Upper
164 100
162 100.9 81.4 122.0
160 43.7 36.6 54.8





#42

2,4,6-Tribromophenol

Concen: 85.513 ng

RT: 10.792 min Scan# 1481

Delta R.T. -0.006 min

Lab File: BF126617.D


Acq: 20 Jan 2022 18:06

Instrument :

BNA_F

ClientSampleId :

MTR-SOIL-6






Tgt Ion:330 Resp: 189579

Ion Ratio Lower Upper

330 100

332 95.8 76.0 114.0

141 37.6 30.1 45.1



#45

2-Fluorobiphenyl

Concen: 61.918 ng

RT: 9.322 min Scan# 1230

Delta R.T. -0.006 min

Lab File: BF126617.D

Acq: 20 Jan 2022 18:06




Tgt Ion:172 Resp: 957752


Ion Ratio Lower Upper

172 100

171 34.9 28.3 42.5

170 23.6 19.0 28.6





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.498 min Scan# 1

Delta R.T. -0.000 min

Lab File: BF126617.D

Acq: 20 Jan 2022 18:06

Instrument :

BNA_F

ClientSampleId :

MTR-SOIL-6





Tgt Ion:188 Resp: 429926

Ion Ratio Lower Upper

188 100

94 11.8 10.1 15.1

80 13.7 11.8 17.6



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.145 min Scan# 2050

Delta R.T. -0.006 min

Lab File: BF126617.D

Acq: 20 Jan 2022 18:06




Tgt Ion:240 Resp: 307861


Ion Ratio Lower Upper

240 100

120 10.5 14.4 21.6#

236 25.4 20.5 30.7









#79
 Terphenyl-d14
 Concen: 60.178 ng
 RT: 13.086 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BF126617.D
 Acq: 20 Jan 2022 18:06

Instrument :
 BNA_F
 ClientSampleId :
 MTR-SOIL-6

Tgt Ion:244 Resp: 1104563
 Ion Ratio Lower Upper
 244 100
 212 7.4 5.0 7.4#
 122 11.3 13.3 19.9#



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.674 min Scan# 2310
 Delta R.T. -0.000 min
 Lab File: BF126617.D
 Acq: 20 Jan 2022 18:06

Tgt Ion:264 Resp: 218665
 Ion Ratio Lower Upper
 264 100
 260 24.7 19.2 28.8
 265 20.6 17.4 26.2

