



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012022\
 Data File : BF126605.D
 Acq On : 20 Jan 2022 11:34
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jan 19 07:04:14 2022






AutoFind: Scans 1641, 1642, 1643; Background Corrected with Scan 1635

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.5	83219	PASS
68	69	0.00	2	1.7	1806	PASS
69	198	0.00	100	50.6	103955	PASS
70	69	0.00	2	0.6	572	PASS
127	198	10	80	50.1	102944	PASS
197	198	0.00	2	0.7	1345	PASS
198	198	100	100	100.0	205411	PASS
199	198	5	9	6.5	13299	PASS
275	198	10	60	26.5	54389	PASS
365	198	1	100	3.7	7512	PASS
441	198	0.01	100	13.8	28360	PASS
442	442	50	100	100.0	175885	PASS
443	442	15	24	18.9	33245	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012022\
Data File : BF126605.D
Acq On : 20 Jan 2022 11:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jan 20 15:37:41 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 19 07:04:14 2022
Response via : Initial Calibration



TIC: BF126605.D\data.ms

(70) Pentachlorophenol (C)

11.292min (-0.000) 29477.15 ng




response 210771

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.40	63.91
264.00	63.00	63.86
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012022\
Data File : BF126605.D
Acq On : 20 Jan 2022 11:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jan 20 15:37:41 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 19 07:04:14 2022
Response via : Initial Calibration



TIC: BF126605.D\data.ms

(77) Benzidine

12.845min (+ 0.012) 0.00 ng

response 1190678

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.80	14.09
183.00	10.20	11.48
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/20/2022	BNA_F	BF126605.D
Compound Name	Response	Retention Time
DDT	509722	13.66
DDD	2659	13.35
DDE	112	13.02
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
2771	512493	0.54

Instrument :
BNA_F
ClientSampleId :
DFTPP