

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012023\
 Data File : BF132155.D
 Acq On : 20 Jan 2023 10:31
 Operator : CG\JU
 Sample : PB150196BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB150196BL

Quant Time: Jan 21 02:19:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 20 00:50:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.089	152	181008	20.000	ng	0.00
21) Naphthalene-d8	8.378	136	656237	20.000	ng	0.00
39) Acenaphthene-d10	10.142	164	363078	20.000	ng	0.00
64) Phenanthrene-d10	11.642	188	736598	20.000	ng	0.00
76) Chrysene-d12	14.301	240	550722	20.000	ng	0.00
86) Perylene-d12	15.901	264	399542	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.725	112	1291387	124.630	ng	0.02
7) Phenol-d6	6.701	99	1682279	128.974	ng	0.00
23) Nitrobenzene-d5	7.654	82	1113228	90.117	ng	0.00
42) 2,4,6-Tribromophenol	10.936	330	523745	135.048	ng	0.00
45) 2-Fluorobiphenyl	9.454	172	1859390	74.832	ng	0.00
79) Terphenyl-d14	13.242	244	2431647	73.962	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012023\
Data File : BF132155.D
Acq On : 20 Jan 2023 10:31
Operator : CG\JU
Sample : PB150196BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB150196BL

Quant Time: Jan 21 02:19:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 20 00:50:42 2023
Response via : Initial Calibration

