

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012023\
 Data File : BF132173.D
 Acq On : 20 Jan 2023 21:41
 Operator : CG\JU
 Sample : 01214-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TJRC-011723-B3-10-20

Quant Time: Jan 21 03:57:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 20 00:50:42 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 01/23/2023
 Supervised By :Jagrut Upadhyay 01/23/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.089	152	170704	20.000	ng	0.00
21) Naphthalene-d8	8.378	136	623011	20.000	ng	0.00
39) Acenaphthene-d10	10.142	164	339132	20.000	ng	0.00
64) Phenanthrene-d10	11.636	188	668403	20.000	ng	0.00
76) Chrysene-d12	14.301	240	447513	20.000	ng	0.00
86) Perylene-d12	15.895	264	369959	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.725	112	1171155	119.849	ng	0.02
7) Phenol-d6	6.701	99	1441183	117.159	ng	0.00
23) Nitrobenzene-d5	7.654	82	1099286	93.734	ng	0.00
42) 2,4,6-Tribromophenol	10.936	330	533172	147.186	ng	0.00
45) 2-Fluorobiphenyl	9.454	172	1810996	78.031	ng	0.00
79) Terphenyl-d14	13.236	244	2290246	85.727	ng	0.00
Target Compounds						
32) Benzoic acid	8.054	122	39031m	12.778	ng	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012023\
Data File : BF132173.D
Acq On : 20 Jan 2023 21:41
Operator : CG\JU
Sample : 01214-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TJRC-011723-B3-10-20

Quant Time: Jan 21 03:57:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 20 00:50:42 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 01/23/2023
Supervised By :Jagrut Upadhyay 01/23/2023

