

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012023\
 Data File : BF132176.D
 Acq On : 20 Jan 2023 23:19
 Operator : CG\JU
 Sample : 01232-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-P-17B

Quant Time: Jan 21 03:58:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 20 00:50:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ng	-7.09
21) Naphthalene-d8	0.000	136	0	0.000	ng	-8.38
39) Acenaphthene-d10	0.000	164	0	0.000	ng	-10.14
64) Phenanthrene-d10	11.636	188	151	20.000	ng	# 0.00
76) Chrysene-d12	0.000	240	0	0.000	ng	-14.31
86) Perylene-d12	0.000	264	0	0.000	ng	-15.90
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0	0.000	ng	
7) Phenol-d6	0.000	99	0	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0	0.000	ng	
79) Terphenyl-d14	13.224	244	662	0.000	ng	-0.01
Target Compounds						
						Qvalue
71) Phenanthrene	11.654	178	307	38.742	ng	# 60
72) Anthracene	11.713	178	151	18.836	ng	# 60
74) Di-n-butylphthalate	12.189	149	1306	171.323	ng	# 78
75) Fluoranthene	12.860	202	361	44.350	ng	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012023\
Data File : BF132176.D
Acq On : 20 Jan 2023 23:19
Operator : CG\JU
Sample : 01232-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-P-17B

Quant Time: Jan 21 03:58:29 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 20 00:50:42 2023
Response via : Initial Calibration

