

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012115\
 Data File : BF077001.D
 Acq On : 21 Jan 2015 18:07
 Operator : TP/IZ
 Sample : G1100-12DL 5X
 Misc : W/ACID
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 8035DL

Quant Time: Jan 22 01:25:00 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF011415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 22 00:29:07 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	34091	20.00	ng	0.00
21) Naphthalene-d8	10.76	136	148143	20.00	ng	-0.01
38) Acenaphthene-d10	14.88	164	75410	20.00	ng	-0.01
63) Phenanthrene-d10	17.69	188	125620	20.00	ng	0.00
75) Chrysene-d12	21.19	240	132168	20.00	ng	-0.01
86) Perylene-d12	22.85	264	115124	20.00	ng	-0.06
System Monitoring Compounds						
5) 2-Fluorophenol	4.90	112	56109	27.06	ng	0.00
7) Phenol-d6	7.26	99	73974	28.28	ng	-0.02
23) Nitrobenzene-d5	9.14	82	44914	16.80	ng	-0.01
41) 2,4,6-Tribromophenol	16.59	330	17415	28.45	ng	-0.01
44) 2-Fluorobiphenyl	13.42	172	92726	18.71	ng	0.00
78) Terphenyl-d14	19.92	244	102981	17.94	ng	-0.01
Target Compounds						Qvalue
8) 2-Chlorophenol	7.38	128	63159	26.42	ng	99
10) Phenol	7.28	94	55893	20.12	ng	88
17) 2-Methylphenol	8.64	107	31204	15.97	ng	# 89
20) 3+4-Methylphenols	9.02	107	86066	33.27	ng	94
26) 2-Nitrophenol	9.93	139	14064	11.48	ng	# 89
27) 2,4-Dimethylphenol	10.22	122	51597	24.49	ng	96
29) 2,4-Dichlorophenol	10.54	162	25310	12.89	ng	94
36) 4-Chloro-3-methylphenol	12.35	107	38790	17.26	ng	97
42) 2,4,6-Trichlorophenol	13.21	196	40851	30.07	ng	93
43) 2,4,5-Trichlorophenol	13.29	196	38173	27.55	ng	98
55) 4-Nitrophenol	15.53	139	23178	27.65	ng	93
64) 4,6-Dinitro-2-methylphenol	16.33	198	9945	14.78	ng	# 65
69) Pentachlorophenol	17.44	266	8641	19.75	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012115\
Data File : BF077001.D
Acq On : 21 Jan 2015 18:07
Operator : TP/IZ
Sample : G1100-12DL 5X
Misc : W/ACID
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
8035DL

Quant Time: Jan 22 01:25:00 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF011415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 22 00:29:07 2015
Response via : Initial Calibration

