

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012215\
 Data File : BF077033.D
 Acq On : 23 Jan 2015 5:42
 Operator : TP/IZ
 Sample : G1149-01
 Misc : S
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-01-011-A

Quant Time: Jan 23 06:36:34 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF011415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 23 03:40:26 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	23628	20.00	ng	0.00
21) Naphthalene-d8	10.76	136	100655	20.00	ng	-0.01
38) Acenaphthene-d10	14.88	164	51955	20.00	ng	-0.01
63) Phenanthrene-d10	17.69	188	87378	20.00	ng	0.00
75) Chrysene-d12	21.19	240	86054	20.00	ng	0.00
86) Perylene-d12	22.85	264	79336	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	4.89	112	125693	87.46	ng	0.00
7) Phenol-d6	7.25	99	164989	91.01	ng	-0.01
23) Nitrobenzene-d5	9.14	82	99864	54.97	ng	-0.01
41) 2,4,6-Tribromophenol	16.59	330	38262	90.73	ng	0.00
44) 2-Fluorobiphenyl	13.41	172	169550	49.66	ng	-0.01
78) Terphenyl-d14	19.92	244	172164	46.06	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	14.48	163	21127	5.73	ng	97
70) Phenanthrene	17.73	178	24737	4.54	ng	96
74) Fluoranthene	19.37	202	31280	5.60	ng	98
77) Pyrene	19.65	202	29438	4.99	ng	96
80) Benzo(a)anthracene	21.18	228	17308	3.21	ng	96
82) Chrysene	21.21	228	11797m	2.40	ng	
87) Benzo(b)fluoranthene	22.44	252	15137m	2.83	ng	
89) Benzo(a)pyrene	22.79	252	10434	2.21	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012215\
Data File : BF077033.D
Acq On : 23 Jan 2015 5:42
Operator : TP/IZ
Sample : G1149-01
Misc : S
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FK-GF-01-011-A

Quant Time: Jan 23 06:36:34 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF011415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 23 03:40:26 2015
Response via : Initial Calibration

