

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012216\
 Data File : BF084590.D
 Acq On : 22 Jan 2016 15:36
 Operator : SJ/IZ
 Sample : H1185-01 5X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GRAB

Manual Integrations
 APPROVED

mohammad
 1/25/2016 2:58:51 PM

Quant Time: Jan 22 23:46:55 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF123115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 14 14:15:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	188442m	20.00	ng	-0.09
21) Naphthalene-d8	8.03	136	813032	20.00	ng	-0.09
38) Acenaphthene-d10	9.79	164	344341	20.00	ng	-0.08
63) Phenanthrene-d10	11.26	188	649663	20.00	ng	-0.09
75) Chrysene-d12	13.89	240	440866	20.00	ng	-0.10
86) Perylene-d12	15.29	264	385129	20.00	ng	-0.13
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	14744	1.27	ng	-0.10
7) Phenol-d6	6.40	99	15241	1.04	ng	-0.08
23) Nitrobenzene-d5	7.31	82	37264	2.55	ng	-0.09
41) 2,4,6-Tribromophenol	10.57	330	4333	1.25	ng	-0.10
44) 2-Fluorobiphenyl	9.10	172	63854	-3.31	ng	-0.09
78) Terphenyl-d14	12.85	244	57377	2.91	ng	-0.09
Target Compounds						
15) Benzyl Alcohol	6.86	79	169064	13.74	ng	# 54
32) Benzoic acid	7.78	122	19756	8.21	ng	97
49) Dimethylphthalate	9.50	163	60822	2.47	ng	# 94
76) Benzidine	12.50	184	79749	5.05	ng	# 88
83) Bis(2-ethylhexyl)phthalate	13.89	149	89538	4.73	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012216\
Data File : BF084590.D
Acq On : 22 Jan 2016 15:36
Operator : SJ/IZ
Sample : H1185-01 5X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GRAB

Manual Integrations
APPROVED

mohammad
1/25/2016 2:58:51 PM

Quant Time: Jan 22 23:46:55 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF123115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 14 14:15:52 2016
Response via : Initial Calibration

