




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012225\
Data File : BF141235.D
Acq On : 22 Jan 2025 13:58
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 22 15:09:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141235.D\data.ms

(70) Pentachlorophenol (C)




11.134min (-0.012) 56694.73 ng

response 213582

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	65.12
264.00	62.70	63.99
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012225\
Data File : BF141235.D
Acq On : 22 Jan 2025 13:58
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 22 15:09:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141235.D\data.ms

(77) Benzidine**12.669min (-0.012) 98791.12 ng****response 1011275**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.34
183.00	11.50	12.09
0.00	0.00	0.00