


Data Path : S:\HPCHEM1\BNA_F\DATA\BF012315\
 Data File : BF077036.D
 Acq On : 23 Jan 2015 11:42
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 23 15:21:24 2015



AutoFind: Scans 1500, 1501, 1502; Background Corrected with Scan 1497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.2	4016	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	42.3	4828	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.4	5863	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	11402	PASS
199	198	5	9	5.5	630	PASS
275	198	10	60	21.0	2394	PASS
365	198	1	100	1.6	182	PASS
441	442	0.01	24	16.9	1786	PASS
442	198	50	100	92.5	10546	PASS
443	442	15	24	18.1	1907	PASS