

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_F\DATA\BF012318\
 Data File : BF102386.D
 Acq On : 24 Jan 2018 12:26
 Operator : SJ/JU
 Sample : J1214-02DL2 20X
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-PH3-MGP-02DL2

Manual Integrations
 APPROVED

Sohil
 1/24/2018 2:47:45 PM

Quant Time: Jan 24 13:31:13 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF012218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 23 00:43:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.71	152	133148	20.00	ng	0.00
21) Naphthalene-d8	8.00	136	545677	20.00	ng	0.00
38) Acenaphthene-d10	9.75	164	231035	20.00	ng	0.00
63) Phenanthrene-d10	11.23	188	334840	20.00	ng	0.00
75) Chrysene-d12	13.87	240	237519	20.00	ng	0.00
86) Perylene-d12	15.29	264	243700	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.32	112	48848	5.56	ng	0.00
7) Phenol-d6	6.35	99	60520	5.72	ng	0.00
23) Nitrobenzene-d5	7.27	82	33556	3.53	ng	-0.01
41) 2,4,6-Tribromophenol	10.54	330	8450	3.69	ng	0.00
44) 2-Fluorobiphenyl	9.07	172	41296	2.63	ng	-0.01
78) Terphenyl-d14	12.82	244	25885	2.46	ng	0.00

Target Compounds

						Qvalue
31) Naphthalene	8.02	128	464162	17.037	ng	100
37) 2-Methylnaphthalene	8.71	142	93717	5.165	ng	97
45) 1,1'-Biphenyl	9.17	154	65928	3.191	ng	98
48) Acenaphthylene	9.61	152	274146	10.959	ng	99
51) Acenaphthene	9.78	154	60798	4.161	ng	99
57) Fluorene	10.29	166	210138	13.412	ng	99
70) Phenanthrene	11.26	178	1055694	54.105	ng	98
71) Anthracene	11.31	178	298635	14.995	ng	99
74) Fluoranthene	12.45	202	477905	24.019	ng	98
77) Pyrene	12.67	202	614137	33.442	ng	99
80) Benzo(a)anthracene	13.86	228	155137m	10.609	ng	
82) Chrysene	13.90	228	151662	10.213	ng	98
85) Indeno(1,2,3-cd)pyrene	16.67	276	50697	4.134	ng	99
87) Benzo(b)fluoranthene	14.89	252	134379m	8.507	ng	
88) Benzo(k)fluoranthene	14.91	252	39131m	2.622	ng	
89) Benzo(a)pyrene	15.23	252	132654	9.312	ng	99
91) Benzo(g,h,i)perylene	17.09	276	59178	5.194	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA_F\DATA\BF012318\
Data File : BF102386.D
Acq On : 24 Jan 2018 12:26
Operator : SJ/JU
Sample : J1214-02DL2 20X
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-PH3-MGP-02DL2

Manual Integrations
APPROVED

Sohil
1/24/2018 2:47:45 PM

Quant Time: Jan 24 13:31:13 2018
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF012218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 23 00:43:41 2018
Response via : Initial Calibration

