

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_F\Data\BF012318\
 Data File : BF102334.D
 Acq On : 23 Jan 2018 9:56
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 23 17:14:53 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF012218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 23 00:43:41 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	39.759	0.6	100	0.01
3	Pyridine	40.000	40.875	-2.2	100	0.01
4	n-Nitrosodimethylamine	40.000	40.948	-2.4	100	0.01
5 S	2-Fluorophenol	80.000	80.482	-0.6	101	0.00
6	Aniline	40.000	39.689	0.8	100	0.00
7 S	Phenol-d6	80.000	79.991	0.0	101	0.01
8	2-Chlorophenol	40.000	41.319	-3.3	103	0.00
9	Benzaldehyde	40.000	40.436	-1.1	101	0.00
10 C	Phenol	40.000	38.932	2.7	98	0.00
11	bis(2-Chloroethyl)ether	40.000	41.045	-2.6	102	0.00
12	1,3-Dichlorobenzene	40.000	40.260	-0.6	102	0.00
13 C	1,4-Dichlorobenzene	40.000	40.286	-0.7	102	0.00
14	1,2-Dichlorobenzene	40.000	41.116	-2.8	103	0.00
15	Benzyl Alcohol	40.000	40.135	-0.3	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.313	-0.8	100	0.00
17	2-Methylphenol	40.000	40.182	-0.5	99	0.00
18	Hexachloroethane	40.000	41.311	-3.3	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.432	-1.1	102	0.00
20	3+4-Methylphenols	40.000	41.157	-2.9	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	39.076	2.3	103	0.00
23 S	Nitrobenzene-d5	80.000	79.245	0.9	102	0.00
24	Nitrobenzene	40.000	39.720	0.7	101	0.00
25	Isophorone	40.000	40.308	-0.8	104	0.00
26 C	2-Nitrophenol	40.000	41.235	-3.1	103	0.00
27	2,4-Dimethylphenol	40.000	39.835	0.4	102	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.117	-0.3	103	0.00
29 C	2,4-Dichlorophenol	40.000	40.183	-0.5	103	0.00
30	1,2,4-Trichlorobenzene	40.000	39.725	0.7	103	0.00
31	Naphthalene	40.000	39.410	1.5	103	0.00
32	Benzoic acid	40.000	36.282	9.3	92	0.01
33	4-Chloroaniline	40.000	39.777	0.6	103	0.00
34 C	Hexachlorobutadiene	40.000	40.658	-1.6	103	0.00
35	Caprolactam	40.000	39.581	1.0	100	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.038	-0.1	102	0.00
37	2-Methylnaphthalene	40.000	39.018	2.5	102	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	40.087	-0.2	103	0.00
40 P	Hexachlorocyclopentadiene	40.000	41.209	-3.0	99	0.00
41 S	2,4,6-Tribromophenol	80.000	77.826	2.7	101	0.00
42 C	2,4,6-Trichlorophenol	40.000	40.481	-1.2	103	0.00
43	2,4,5-Trichlorophenol	40.000	41.219	-3.0	104	0.00
44 S	2-Fluorobiphenyl	80.000	79.942	0.1	102	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_F\Data\BF012318\
 Data File : BF102334.D
 Acq On : 23 Jan 2018 9:56
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 23 17:14:53 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF012218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 23 00:43:41 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	39.878	0.3	104	0.00
46	2-Chloronaphthalene	40.000	39.565	1.1	102	0.00
47	2-Nitroaniline	40.000	41.261	-3.2	106	0.00
48	Acenaphthylene	40.000	39.176	2.1	101	0.00
49	Dimethylphthalate	40.000	39.952	0.1	104	0.00
50	2,6-Dinitrotoluene	40.000	39.690	0.8	99	0.00
51 C	Acenaphthene	40.000	39.357	1.6	103	0.00
52	3-Nitroaniline	40.000	39.556	1.1	101	0.00
53 P	2,4-Dinitrophenol	40.000	39.427	1.4	98	0.00
54	Dibenzofuran	40.000	40.695	-1.7	102	0.00
55 P	4-Nitrophenol	40.000	39.440	1.4	95	0.01
56	2,4-Dinitrotoluene	40.000	39.556	1.1	100	0.00
57	Fluorene	40.000	41.010	-2.5	105	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	40.623	-1.6	106	0.00
59	Diethylphthalate	40.000	39.581	1.0	103	0.00
60	4-Chlorophenyl-phenylether	40.000	40.951	-2.4	103	0.00
61	4-Nitroaniline	40.000	39.943	0.1	102	0.00
62	Azobenzene	40.000	39.582	1.0	102	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
64	4,6-Dinitro-2-methylphenol	40.000	43.246	-8.1	103	0.01
65 c	n-Nitrosodiphenylamine	40.000	40.610	-1.5	103	0.00
66	4-Bromophenyl-phenylether	40.000	41.733	-4.3	104	0.00
67	Hexachlorobenzene	40.000	40.821	-2.1	102	0.00
68	Atrazine	40.000	40.374	-0.9	102	0.00
69 C	Pentachlorophenol	40.000	40.077	-0.2	92	0.00
70	Phenanthrene	40.000	39.477	1.3	101	0.00
71	Anthracene	40.000	39.670	0.8	101	0.00
72	Carbazole	40.000	39.510	1.2	101	0.00
73	Di-n-butylphthalate	40.000	40.053	-0.1	100	0.00
74 C	Fluoranthene	40.000	39.255	1.9	100	0.01
75 I	Chrysene-d12	20.000	20.000	0.0	99	0.01
76	Benzidine	40.000	37.328	6.7	99	0.00
77	Pyrene	40.000	39.981	0.0	99	0.01
78 S	Terphenyl-d14	80.000	78.420	2.0	100	0.00
79	Butylbenzylphthalate	40.000	41.103	-2.8	99	0.00
80	Benzo(a)anthracene	40.000	39.738	0.7	100	0.01
81	3,3'-Dichlorobenzidine	40.000	41.164	-2.9	101	0.00
82	Chrysene	40.000	39.429	1.4	99	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	41.236	-3.1	100	0.00
84 c	Di-n-octyl phthalate	40.000	40.754	-1.9	100	0.01
85	Indeno(1,2,3-cd)pyrene	40.000	38.369	4.1	102	0.02
86 I	Perylene-d12	20.000	20.000	0.0	96	0.01
87	Benzo(b)fluoranthene	40.000	42.239	-5.6	99	0.01

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_F\Data\BF012318\
Data File : BF102334.D
Acq On : 23 Jan 2018 9:56
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 23 17:14:53 2018
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF012218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 23 00:43:41 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	38.092	4.8	95	0.01
89 C	Benzo(a)pyrene	40.000	40.341	-0.9	97	0.01
90	Dibenzo(a,h)anthracene	40.000	39.399	1.5	100	0.02
91	Benzo(g,h,i)perylene	40.000	40.081	-0.2	102	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0