

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012319\
 Data File : BF112202.D
 Acq On : 23 Jan 2019 15:24
 Operator : JU/SJ
 Sample : K1216-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MULCH-1

Manual Integrations
 APPROVED

Sohil
 1/24/2019 11:30:24 AM

Quant Time: Jan 23 16:01:31 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 19 01:02:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	199186	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	792517	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	412074	20.00	ng	-0.01
64) Phenanthrene-d10	11.37	188	827256	20.00	ng	-0.01
76) Chrysene-d12	14.02	240	671412	20.00	ng	-0.01
87) Perylene-d12	15.49	264	557815	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	1094405	99.73	ng	0.00
7) Phenol-d6	6.50	99	1414623	103.03	ng	0.00
23) Nitrobenzene-d5	7.41	82	887653	77.38	ng	-0.02
42) 2,4,6-Tribromophenol	10.68	330	509590	114.63	ng	-0.01
45) 2-Fluorobiphenyl	9.21	172	1804731	64.21	ng	0.00
79) Terphenyl-d14	12.96	244	2089956	66.22	ng	0.00
Target Compounds						
10) Phenol	6.51	94	89197	5.944	ng	Qvalue # 72
50) Dimethylphthalate	9.60	163	159285	5.300	ng	99
71) Phenanthrene	11.40	178	130850	3.143	ng	95
75) Fluoranthene	12.59	202	180922	4.116	ng	99
78) Pyrene	12.82	202	180786	4.081	ng	96
83) Chrysene	14.04	228	80071	2.198	ng	98
88) Benzo(b)fluoranthene	15.06	252	74920m	2.366	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012319\
Data File : BF112202.D
Acq On : 23 Jan 2019 15:24
Operator : JU/SJ
Sample : K1216-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MULCH-1

Manual Integrations
APPROVED

Sohil
1/24/2019 11:30:24 AM

Quant Time: Jan 23 16:01:31 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 19 01:02:12 2019
Response via : Initial Calibration

