

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012320\
 Data File : BF118677.D
 Acq On : 23 Jan 2020 15:17
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 24 02:41:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 20 14:45:03 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.522	0.517	1.0	75	-0.02
3	Pyridine	1.470	1.466	0.3	75	-0.02
4	n-Nitrosodimethylamine	0.630	0.539	14.4	64	-0.02
5 S	2-Fluorophenol	1.077	1.093	-1.5	76	0.00
6	Aniline	1.843	1.846	-0.2	75	0.00
7 S	Phenol-d6	1.400	1.396	0.3	74	0.00
8	2-Chlorophenol	1.209	1.215	-0.5	75	0.00
9	Benzaldehyde	0.851	0.847	0.5	77	0.00
10 C	Phenol	1.595	1.549	2.9	73	0.00
11	bis(2-Chloroethyl)ether	1.183	1.188	-0.4	76	0.00
12	1,3-Dichlorobenzene	1.415	1.469	-3.8	78	0.00
13 C	1,4-Dichlorobenzene	1.420	1.471	-3.6	77	0.00
14	1,2-Dichlorobenzene	1.335	1.372	-2.8	76	0.00
15	Benzyl Alcohol	1.110	1.084	2.3	74	0.00
16	2,2'-oxybis(1-Chloropropane	1.648	1.542	6.4	70	-0.01
17	2-Methylphenol	1.007	1.010	-0.3	76	0.00
18	Hexachloroethane	0.515	0.499	3.1	73	0.00
19 P	n-Nitroso-di-n-propylamine	0.945	0.851	9.9	69	0.00
20	3+4-Methylphenols	1.229	1.204	2.0	72	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
22	Acetophenone	0.474	0.460	3.0	72	0.00
23 S	Nitrobenzene-d5	0.358	0.332	7.3	70	0.00
24	Nitrobenzene	0.366	0.341	6.8	70	0.00
25	Isophorone	0.652	0.629	3.5	74	0.00
26 C	2-Nitrophenol	0.156	0.148	5.1	73	0.00
27	2,4-Dimethylphenol	0.270	0.254	5.9	71	0.00
28	bis(2-Chloroethoxy)methane	0.420	0.414	1.4	75	-0.01
29 C	2,4-Dichlorophenol	0.298	0.302	-1.3	76	0.00
30	1,2,4-Trichlorobenzene	0.346	0.354	-2.3	77	0.00
31	Naphthalene	0.903	0.960	-6.3	78	0.00
32	Benzoic acid	0.202	0.176	12.9	67	0.00
33	4-Chloroaniline	0.418	0.434	-3.8	77	0.00
34 C	Hexachlorobutadiene	0.210	0.212	-1.0	76	-0.01
35	Caprolactam	0.098	0.100	-2.0	81	0.00
36 C	4-Chloro-3-methylphenol	0.294	0.297	-1.0	76	0.00
37	2-Methylnaphthalene	0.638	0.665	-4.2	77	0.00
38	1-Methylnaphthalene	0.606	0.622	-2.6	76	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	79	-0.01
40	1,2,4,5-Tetrachlorobenzene	0.626	0.635	-1.4	75	0.00
41 P	Hexachlorocyclopentadiene	0.316	0.253	19.9	60	0.00
42 S	2,4,6-Tribromophenol	0.231	0.228	1.3	74	0.00
43 C	2,4,6-Trichlorophenol	0.415	0.404	2.7	74	0.00
44	2,4,5-Trichlorophenol	0.424	0.426	-0.5	76	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012320\
 Data File : BF118677.D
 Acq On : 23 Jan 2020 15:17
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 24 02:41:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 20 14:45:03 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.217	1.144	6.0	76	0.00
46	1,1'-Biphenyl	1.378	1.459	-5.9	77	0.00
47	2-Chloronaphthalene	1.162	1.216	-4.6	77	0.00
48	2-Nitroaniline	0.357	0.320	10.4	68	0.00
49	Acenaphthylene	1.634	1.750	-7.1	79	0.00
50	Dimethylphthalate	1.303	1.376	-5.6	80	0.00
51	2,6-Dinitrotoluene	0.292	0.287	1.7	75	0.00
52 C	Acenaphthene	1.093	1.118	-2.3	76	0.00
53	3-Nitroaniline	0.333	0.330	0.9	75	0.00
54 P	2,4-Dinitrophenol	0.116	0.111	4.3	78	0.00
55	Dibenzofuran	1.515	1.599	-5.5	77	0.00
56 P	4-Nitrophenol	0.252	0.246	2.4	74	0.00
57	2,4-Dinitrotoluene	0.368	0.369	-0.3	78	0.00
58	Fluorene	1.189	1.211	-1.9	75	-0.01
59	2,3,4,6-Tetrachlorophenol	0.373	0.373	0.0	75	0.00
60	Diethylphthalate	1.263	1.314	-4.0	78	-0.01
61	4-Chlorophenyl-phenylether	0.647	0.652	-0.8	73	0.00
62	4-Nitroaniline	0.342	0.346	-1.2	78	-0.01
63	Azobenzene	1.224	1.213	0.9	74	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	82	-0.01
65	4,6-Dinitro-2-methylphenol	0.088	0.088	0.0	82	0.00
66 c	n-Nitrosodiphenylamine	0.603	0.602	0.2	77	0.00
67	4-Bromophenyl-phenylether	0.247	0.246	0.4	78	-0.01
68	Hexachlorobenzene	0.280	0.279	0.4	78	0.00
69	Atrazine	0.209	0.207	1.0	78	0.00
70 C	Pentachlorophenol	0.156	0.148	5.1	75	0.00
71	Phenanthrene	0.965	1.002	-3.8	79	0.00
72	Anthracene	0.955	0.996	-4.3	79	0.00
73	Carbazole	0.876	0.913	-4.2	79	0.00
74	Di-n-butylphthalate	0.973	1.042	-7.1	82	-0.01
75 C	Fluoranthene	1.018	1.086	-6.7	80	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
77	Benzidine	0.534	0.653	-22.3	93	0.00
78	Pyrene	1.358	1.404	-3.4	80	0.00
79 S	Terphenyl-d14	0.917	0.935	-2.0	79	0.00
80	Butylbenzylphthalate	0.556	0.582	-4.7	83	-0.01
81	Benzo(a)anthracene	1.203	1.246	-3.6	79	-0.01
82	3,3'-Dichlorobenzidine	0.429	0.498	-16.1	87	-0.01
83	Chrysene	1.153	1.213	-5.2	81	0.00
84	Bis(2-ethylhexyl)phthalate	0.679	0.728	-7.2	84	-0.02
85 c	Di-n-octyl phthalate	1.005	1.177	-17.1	94	-0.02
86	Indeno(1,2,3-cd)pyrene	1.002	1.144	-14.2	99	0.00
87 I	Perylene-d12	1.000	1.000	0.0	100	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012320\
Data File : BF118677.D
Acq On : 23 Jan 2020 15:17
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 24 02:41:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 20 14:45:03 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.257	1.220	2.9	89	0.00
89	Benzo(k)fluoranthene	1.165	1.186	-1.8	98	0.00
90 C	Benzo(a)pyrene	1.084	1.102	-1.7	97	0.00
91	Dibenzo(a,h)anthracene	0.959	0.924	3.6	96	0.00
92	Benzo(q,h,i)perylene	0.931	0.901	3.2	97	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0