




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012325\
Data File : BF141243.D
Acq On : 23 Jan 2025 12:05
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 23 13:22:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration






TIC: BF141243.D\data.ms

(70) Pentachlorophenol (C)**11.134min (-0.012) 29408.38 ng****response 265988**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	62.61
264.00	62.70	63.33
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012325\
Data File : BF141243.D
Acq On : 23 Jan 2025 12:05
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 23 13:22:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141243.D\data.ms

(77) Benzidine

12.675min (-0.006) 67921.69 ng

response 1180213

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.29
183.00	11.50	12.33
0.00	0.00	0.00