

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012417\
 Data File : BF092432.D
 Acq On : 24 Jan 2017 16:47
 Operator : UM/SJ
 Sample : I1304-07
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ROLL-OFF-15939-COMP

Manual Integrations
 APPROVED

mohammad
 1/25/2017 3:49:36 PM

Quant Time: Jan 24 17:15:42 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF012417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 24 13:48:07 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	152639	20.00	ng	0.00
21) Naphthalene-d8	8.26	136	605788	20.00	ng	-0.01
38) Acenaphthene-d10	10.03	164	295650	20.00	ng	0.00
63) Phenanthrene-d10	11.52	188	465639	20.00	ng	-0.01
75) Chrysene-d12	14.16	240	339727	20.00	ng	-0.01
86) Perylene-d12	15.63	264	302978	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.57	112	1347230	137.32	ng	0.02
7) Phenol-d6	6.59	99	1498011	119.87	ng	0.00
23) Nitrobenzene-d5	7.53	82	1026803	92.59	ng	-0.01
41) 2,4,6-Tribromophenol	10.82	330	256536	127.33	ng	0.00
44) 2-Fluorobiphenyl	9.34	172	1433699	89.67	ng	-0.01
78) Terphenyl-d14	13.11	244	1051334	82.41	ng	-0.01
Target Compounds						
10) Phenol	6.60	94	76931	5.10	ng	# 70
49) Dimethylphthalate	9.73	163	213745	11.37	ng	# 97
70) Phenanthrene	11.54	178	87978	3.43	ng	98
74) Fluoranthene	12.73	202	156264	6.23	ng	98
77) Pyrene	12.96	202	157479	6.40	ng	97
80) Benzo(a)anthracene	14.15	228	65042	3.56	ng	94
82) Chrysene	14.18	228	74217m	3.81	ng	
83) Bis(2-ethylhexyl)phthalate	14.15	149	36240	2.89	ng	100
85) Indeno(1,2,3-cd)pyrene	17.11	276	35477	2.46	ng	# 85
87) Benzo(b)fluoranthene	15.20	252	70887m	3.64	ng	
88) Benzo(k)fluoranthene	15.21	252	36073m	2.23	ng	
89) Benzo(a)pyrene	15.56	252	56476	3.43	ng	# 94
91) Benzo(a,h,i)perylene	17.55	276	39323	2.92	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012417\
Data File : BF092432.D
Acq On : 24 Jan 2017 16:47
Operator : UM/SJ
Sample : I1304-07
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ROLL-OFF-15939-COMP

Manual Integrations
APPROVED

mohammad
1/25/2017 3:49:36 PM

Quant Time: Jan 24 17:15:42 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF012417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 24 13:48:07 2017
Response via : Initial Calibration

