

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012419\
 Data File : BF112222.D
 Acq On : 24 Jan 2019 21:24
 Operator : JU/SJ
 Sample : K1223-10 2X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TS-V

Manual Integrations
 APPROVED

Sohil
 1/25/2019 10:05:26 AM

Quant Time: Jan 25 00:56:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 19 01:02:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	198837	20.00	ng	0.00
21) Naphthalene-d8	8.14	136	744119	20.00	ng	0.00
39) Acenaphthene-d10	9.90	164	316098	20.00	ng	0.00
64) Phenanthrene-d10	11.39	188	529337	20.00	ng	0.00
76) Chrysene-d12	14.03	240	359269	20.00	ng	0.00
87) Perylene-d12	15.52	264	269040	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	482299	44.03	ng	0.00
7) Phenol-d6	6.50	99	675128	49.26	ng	0.00
23) Nitrobenzene-d5	7.42	82	390768	36.28	ng	-0.01
42) 2,4,6-Tribromophenol	10.69	330	158831	46.58	ng	0.00
45) 2-Fluorobiphenyl	9.22	172	733564	34.02	ng	0.00
79) Terphenyl-d14	12.97	244	506491	29.99	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.60	163	67331	2.921	ng	Qvalue # 99
71) Phenanthrene	11.41	178	246974	9.271	ng	95
75) Fluoranthene	12.60	202	361173	12.841	ng	97
78) Pyrene	12.83	202	293566	12.385	ng	98
81) Benzo(a)anthracene	14.02	228	136595	6.631	ng	97
83) Chrysene	14.06	228	140305	7.198	ng	97
84) Bis(2-ethylhexyl)phthalate	14.00	149	58205	3.906	ng	98
86) Indeno(1,2,3-cd)pyrene	17.01	276	60193	3.347	ng	95
88) Benzo(b)fluoranthene	15.08	252	157113m	10.288	ng	
89) Benzo(k)fluoranthene	15.10	252	49552m	3.352	ng	
90) Benzo(a)pyrene	15.45	252	93840	6.700	ng	# 86
92) Benzo(a,h,i)perylene	17.46	276	54765	4.469	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012419\
Data File : BF112222.D
Acq On : 24 Jan 2019 21:24
Operator : JU/SJ
Sample : K1223-10 2X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :
TS-V

Quant Time: Jan 25 00:56:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 19 01:02:12 2019
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
1/25/2019 10:05:26 AM

