

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012420\
 Data File : BF118689.D
 Acq On : 24 Jan 2020 16:52
 Operator : CG/JU
 Sample : L1246-05
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-3

Quant Time: Jan 27 01:43:23 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 20 14:45:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	278633	20.00	ng	-0.02
21) Naphthalene-d8	8.15	136	1026773	20.00	ng	-0.02
39) Acenaphthene-d10	9.90	164	564743	20.00	ng	-0.03
64) Phenanthrene-d10	11.38	188	980739	20.00	ng	-0.03
76) Chrysene-d12	14.01	240	614542	20.00	ng	-0.04
87) Perylene-d12	15.47	264	662851	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	1500260	100.00	ng	-0.02
7) Phenol-d6	6.49	99	1854926	95.13	ng	-0.02
23) Nitrobenzene-d5	7.42	82	1188382	64.74	ng	-0.02
42) 2,4,6-Tribromophenol	10.68	330	594636	91.03	ng	-0.03
45) 2-Fluorobiphenyl	9.22	172	2250991	65.51	ng	-0.03
79) Terphenyl-d14	12.96	244	2332355	82.78	ng	-0.03
Target Compounds						
10) Phenol	6.50	94	97193	4.374	ng	95
50) Dimethylphthalate	9.61	163	475337	12.919	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012420\
Data File : BF118689.D
Acq On : 24 Jan 2020 16:52
Operator : CG/JU
Sample : L1246-05
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP-3

Quant Time: Jan 27 01:43:23 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 20 14:45:03 2020
Response via : Initial Calibration

