




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012425\
Data File : BF141255.D
Acq On : 24 Jan 2025 10:46
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 25 00:53:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141255.D\data.ms




(70) Pentachlorophenol (C)

11.134min (-0.012) 29479.71 ng

response 204451

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	63.79
264.00	62.70	63.39
0.00	0.00	0.00

Quant Time: Jan 25 00:53:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141255.D\data.ms

(77) Benzidine

12.669min (-0.012) 62741.67 ng

response **766750**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	13.96
183.00	11.50	12.09
0.00	0.00	0.00