

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012516\  
 Data File : BF084613.D  
 Acq On : 25 Jan 2016 17:59  
 Operator : SJ/IZ  
 Sample : H1201-08  
 Misc :  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 MW-105

Quant Time: Jan 26 09:41:27 2016  
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF123115.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jan 25 16:24:07 2016  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.74  | 152  | 194246   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.02  | 136  | 774234   | 20.00  | ng    | -0.01    |
| 38) Acenaphthene-d10        | 9.77  | 164  | 367354   | 20.00  | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.25 | 188  | 777656   | 20.00  | ng    | -0.01    |
| 75) Chrysene-d12            | 13.88 | 240  | 654991   | 20.00  | ng    | -0.01    |
| 86) Perylene-d12            | 15.28 | 264  | 398354   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.33  | 112  | 780805   | 65.02  | ng    | 0.00     |
| 7) Phenol-d6                | 6.37  | 99   | 618261   | 40.99  | ng    | -0.02    |
| 23) Nitrobenzene-d5         | 7.31  | 82   | 1373937  | 98.76  | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 10.57 | 330  | 621906   | 167.69 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 9.10  | 172  | 2138630  | 103.81 | ng    | 0.00     |
| 78) Terphenyl-d14           | 12.84 | 244  | 1937253  | 66.02  | ng    | -0.01    |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012516\  
Data File : BF084613.D  
Acq On : 25 Jan 2016 17:59  
Operator : SJ/IZ  
Sample : H1201-08  
Misc :  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
MW-105

Quant Time: Jan 26 09:41:27 2016  
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF123115.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jan 25 16:24:07 2016  
Response via : Initial Calibration

