

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012524\
 Data File : BF137148.D
 Acq On : 25 Jan 2024 15:31
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF012524

Quant Time: Jan 25 23:15:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 25 23:14:32 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	40.758	-1.9	103	0.00
3	Pyridine	40.000	41.683	-4.2	104	0.00
4	n-Nitrosodimethylamine	40.000	41.538	-3.8	102	0.00
5 S	2-Fluorophenol	80.000	82.664	-3.3	104	0.00
6	Aniline	40.000	41.283	-3.2	103	0.00
7 S	Phenol-d6	80.000	81.937	-2.4	102	0.00
8	2-Chlorophenol	40.000	41.514	-3.8	103	0.00
9	Benzaldehyde	40.000	42.560	-6.4	102	0.00
10 C	Phenol	40.000	40.352	-0.9	101	0.00
11	bis(2-Chloroethyl)ether	40.000	41.033	-2.6	103	0.00
12	1,3-Dichlorobenzene	40.000	40.909	-2.3	103	0.00
13 C	1,4-Dichlorobenzene	40.000	41.207	-3.0	103	0.00
14	1,2-Dichlorobenzene	40.000	41.067	-2.7	102	0.00
15	Benzyl Alcohol	40.000	42.242	-5.6	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.382	1.5	98	0.00
17	2-Methylphenol	40.000	42.367	-5.9	106	0.00
18	Hexachloroethane	40.000	42.422	-6.1	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.345	-0.9	101	0.00
20	3+4-Methylphenols	40.000	41.710	-4.3	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	39.679	0.8	101	0.00
23 S	Nitrobenzene-d5	80.000	88.807	-11.0	105	0.00
24	Nitrobenzene	40.000	43.608	-9.0	104	0.00
25	Isophorone	40.000	40.634	-1.6	103	0.00
26 C	2-Nitrophenol	40.000	45.212	-13.0	122	0.00
27	2,4-Dimethylphenol	40.000	41.279	-3.2	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.450	-1.1	102	0.00
29 C	2,4-Dichlorophenol	40.000	41.958	-4.9	104	0.00
30	1,2,4-Trichlorobenzene	40.000	40.528	-1.3	102	0.00
31	Naphthalene	40.000	40.489	-1.2	103	0.00
32	Benzoic acid	40.000	45.078	-12.7	133	0.00
33	4-Chloroaniline	40.000	41.264	-3.2	104	0.00
34 C	Hexachlorobutadiene	40.000	40.382	-1.0	103	0.00
35	Caprolactam	40.000	43.484	-8.7	109	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.064	-5.2	103	0.00
37	2-Methylnaphthalene	40.000	40.669	-1.7	103	0.00
38	1-Methylnaphthalene	40.000	40.758	-1.9	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.643	0.9	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.679	-14.2	108	0.00
42 S	2,4,6-Tribromophenol	80.000	85.662	-7.1	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.738	-6.8	108	0.00
44	2,4,5-Trichlorophenol	40.000	40.788	-2.0	103	0.00
45 S	2-Fluorobiphenyl	80.000	77.869	2.7	102	0.00
46	1,1'-Biphenyl	40.000	39.799	0.5	103	0.00
47	2-Chloronaphthalene	40.000	40.077	-0.2	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012524\
 Data File : BF137148.D
 Acq On : 25 Jan 2024 15:31
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF012524

Quant Time: Jan 25 23:15:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 25 23:14:32 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.420	-6.1	113	0.00
49	Acenaphthylene	40.000	39.957	0.1	104	0.00
50	Dimethylphthalate	40.000	40.338	-0.8	104	0.00
51	2,6-Dinitrotoluene	40.000	43.446	-8.6	115	0.00
52 C	Acenaphthene	40.000	40.114	-0.3	105	0.00
53	3-Nitroaniline	40.000	47.326	-18.3	114	0.00
54 P	2,4-Dinitrophenol	40.000	46.280	-15.7	134	0.00
55	Dibenzofuran	40.000	39.962	0.1	103	0.00
56 P	4-Nitrophenol	40.000	48.551	-21.4	118	0.00
57	2,4-Dinitrotoluene	40.000	43.890	-9.7	120	0.00
58	Fluorene	40.000	39.073	2.3	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.242	-5.6	103	0.00
60	Diethylphthalate	40.000	40.729	-1.8	103	0.00
61	4-Chlorophenyl-phenylether	40.000	39.581	1.0	102	0.00
62	4-Nitroaniline	40.000	46.749	-16.9	111	0.00
63	Azobenzene	40.000	40.030	-0.1	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.696	-14.2	136	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.117	-0.3	104	0.00
67	4-Bromophenyl-phenylether	40.000	40.851	-2.1	107	0.00
68	Hexachlorobenzene	40.000	40.314	-0.8	105	0.00
69	Atrazine	40.000	43.201	-8.0	110	0.00
70 C	Pentachlorophenol	40.000	45.874	-14.7	112	0.00
71	Phenanthrene	40.000	39.929	0.2	106	0.00
72	Anthracene	40.000	40.105	-0.3	106	0.00
73	Carbazole	40.000	41.308	-3.3	108	0.00
74	Di-n-butylphthalate	40.000	44.151	-10.4	111	0.00
75 C	Fluoranthene	40.000	41.665	-4.2	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	117	0.00
77	Benzydine	40.000	47.163	-17.9	119	0.00
78	Pyrene	40.000	39.443	1.4	110	0.00
79 S	Terphenyl-d14	80.000	78.611	1.7	112	0.00
80	Butylbenzylphthalate	40.000	47.383	-18.5	151	0.00
81	Benzo(a)anthracene	40.000	41.836	-4.6	119	0.00
82	3,3'-Dichlorobenzidine	40.000	47.742	-19.4	124	0.00
83	Chrysene	40.000	41.179	-2.9	118	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.232	-10.6	145	0.00
85 c	Di-n-octyl phthalate	40.000	44.456	-11.1	145	0.00
86 I	Perylene-d12	20.000	20.000	0.0	109	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.009	2.5	97	0.00
88	Benzo(b)fluoranthene	40.000	41.497	-3.7	111	0.00
89	Benzo(k)fluoranthene	40.000	43.329	-8.3	119	0.00
90 C	Benzo(a)pyrene	40.000	41.465	-3.7	110	0.00
91	Dibenzo(a,h)anthracene	40.000	39.382	1.5	98	0.00
92	Benzo(g,h,i)perylene	40.000	40.028	-0.1	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012524\
Data File : BF137148.D
Acq On : 25 Jan 2024 15:31
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF012524

Quant Time: Jan 25 23:15:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 25 23:14:32 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0