

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF012615\
 Data File : BF077089.D
 Acq On : 27 Jan 2015 17:36
 Operator : TP/IZ
 Sample : G1137-07
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 104B1A

Manual Integrations
 APPROVED

apatel
 1/28/2015 5:43:09 PM

Quant Time: Jan 28 15:05:32 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 28 02:59:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.76	152	28106	20.00	ng	0.00
21) Naphthalene-d8	10.68	136	114112	20.00	ng	0.00
38) Acenaphthene-d10	14.80	164	54730	20.00	ng	0.00
63) Phenanthrene-d10	17.64	188	99013	20.00	ng	0.00
75) Chrysene-d12	21.15	240	104384	20.00	ng	-0.01
86) Perylene-d12	22.88	264	83110	20.00	ng	-0.08
System Monitoring Compounds						
5) 2-Fluorophenol	4.80	112	64870	37.95	ng	0.00
7) Phenol-d6	7.20	99	47985	22.25	ng	0.02
23) Nitrobenzene-d5	9.06	82	146099	70.93	ng	0.00
41) 2,4,6-Tribromophenol	16.53	330	46047	103.65	ng	0.00
44) 2-Fluorobiphenyl	13.34	172	287488	79.94	ng	0.01
78) Terphenyl-d14	19.88	244	301356	66.46	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.72	128	29542	5.03	ng	97
37) 2-Methylnaphthalene	12.38	142	20224	5.04	ng	96
49) Dimethylphthalate	14.40	163	17834	4.59	ng	98
51) Acenaphthene	14.87	154	6407	2.01	ng	88
57) Fluorene	16.05	166	11615	2.95	ng	# 94
59) Diethylphthalate	16.05	149	11091	2.83	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF012615\
Data File : BF077089.D
Acq On : 27 Jan 2015 17:36
Operator : TP/IZ
Sample : G1137-07
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
104B1A

Manual Integrations
APPROVED

apatel
1/28/2015 5:43:09 PM

Quant Time: Jan 28 15:05:32 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 28 02:59:18 2015
Response via : Initial Calibration

