

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF012615\
 Data File : BF077091.D
 Acq On : 27 Jan 2015 18:45
 Operator : TP/IZ
 Sample : G1137-09
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 104B2

Quant Time: Jan 28 15:39:21 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 28 02:59:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.76	152	30992	20.00	ng	0.00
21) Naphthalene-d8	10.68	136	133451	20.00	ng	0.00
38) Acenaphthene-d10	14.80	164	62734	20.00	ng	0.00
63) Phenanthrene-d10	17.64	188	110330	20.00	ng	0.00
75) Chrysene-d12	21.15	240	111054	20.00	ng	-0.01
86) Perylene-d12	22.80	264	106568	20.00	ng	-0.16
System Monitoring Compounds						
5) 2-Fluorophenol	4.80	112	72653	38.54	ng	0.00
7) Phenol-d6	7.18	99	55805	23.47	ng	0.00
23) Nitrobenzene-d5	9.06	82	171301	71.11	ng	0.00
41) 2,4,6-Tribromophenol	16.53	330	54328	106.69	ng	0.00
44) 2-Fluorobiphenyl	13.34	172	332955	80.77	ng	0.01
78) Terphenyl-d14	19.88	244	347265	71.99	ng	0.00
Target Compounds						
10) Phenol	7.21	94	5564	2.20	ng	Qvalue # 68
49) Dimethylphthalate	14.40	163	17459	3.92	ng	97
51) Acenaphthene	14.87	154	11585	3.16	ng	87
57) Fluorene	16.05	166	20045	4.43	ng	99
72) Carbazole	18.04	167	56172	8.63	ng	98
73) Di-n-butylphthalate	18.63	149	20318	2.70	ng	Qvalue # 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF012615\
Data File : BF077091.D
Acq On : 27 Jan 2015 18:45
Operator : TP/IZ
Sample : G1137-09
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
104B2

Quant Time: Jan 28 15:39:21 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 28 02:59:18 2015
Response via : Initial Calibration

