

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF012615\
 Data File : BF077081.D
 Acq On : 27 Jan 2015 9:25
 Operator : TP/IZ
 Sample : G1044-03MSD
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD

Manual Integrations
 APPROVED

apatel
 1/28/2015 5:42:50 PM

Quant Time: Jan 28 14:21:22 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 23 15:21:24 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	35587	20.00	ng	0.01
21) Naphthalene-d8	10.69	136	151663	20.00	ng	0.01
38) Acenaphthene-d10	14.81	164	86014	20.00	ng	0.01
63) Phenanthrene-d10	17.64	188	139399	20.00	ng	0.00
75) Chrysene-d12	21.15	240	130400	20.00	ng	0.00
86) Perylene-d12	22.79	264	117008	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	4.82	112	140280	64.81	ng	0.05
7) Phenol-d6	7.21	99	118809	43.51	ng	0.05
23) Nitrobenzene-d5	9.09	82	259879	94.93	ng	0.01
41) 2,4,6-Tribromophenol	16.54	330	100949	144.58	ng	0.01
44) 2-Fluorobiphenyl	13.34	172	523446	92.61	ng	0.00
78) Terphenyl-d14	19.88	244	479268	84.61	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	1.16	88	20453	22.07	ng	92
3) Pyridine	1.64	79	28605	11.93	ng	90
4) n-Nitrosodimethylamine	1.58	42	36632	30.84	ng	81
6) Aniline	7.06	93	164227	43.46	ng	98
8) 2-Chlorophenol	7.32	128	112341	45.02	ng	95
9) Benzaldehyde	6.79	77	7307	3.54	ng	# 1
10) Phenol	7.25	94	87271m	30.10	ng	
11) bis(2-Chloroethyl)ether	7.32	93	132774	58.53	ng	# 85
12) 1,3-Dichlorobenzene	7.61	146	119906	42.24	ng	98
13) 1,4-Dichlorobenzene	7.81	146	158628	52.69	ng	99
14) 1,2-Dichlorobenzene	8.13	146	153418	55.31	ng	99
15) Benzyl Alcohol	8.24	79	65379	29.59	ng	98
16) 2,2'-oxybis(1-Chloropropan	8.56	45	131805	43.22	ng	# 47
17) 2-Methylphenol	8.60	107	78108	38.30	ng	93
18) Hexachloroethane	8.87	117	45374	43.33	ng	98
19) n-Nitroso-di-n-propylamine	8.86	70	87960	46.02	ng	97
20) 3+4-Methylphenols	8.97	107	155053	57.43	ng	98
22) Acetophenone	8.77	105	190099	51.17	ng	# 98
24) Nitrobenzene	9.12	77	136754	49.04	ng	97
25) Isophorone	9.74	82	230797	46.29	ng	95
26) 2-Nitrophenol	9.85	139	63175	44.38	ng	# 82
27) 2,4-Dimethylphenol	10.16	122	117587	54.53	ng	97
28) bis(2-Chloroethoxy)methane	10.36	93	148856	52.53	ng	97
29) 2,4-Dichlorophenol	10.48	162	89215	44.39	ng	96
30) 1,2,4-Trichlorobenzene	10.60	180	99458	44.55	ng	99
31) Naphthalene	10.73	128	706475	90.48	ng	99
32) Benzoic acid	10.61	122	76548m	64.07	ng	
33) 4-Chloroaniline	10.99	127	42594	13.50	ng	92
34) Hexachlorobutadiene	11.10	225	56786	42.87	ng	97
36) 4-Chloro-3-methylphenol	12.32	107	234996	102.11	ng	# 54
37) 2-Methylnaphthalene	12.39	142	288744	54.15	ng	97
39) 1,2,4,5-Tetrachlorobenzene	12.79	216	95135	44.74	ng	96
40) Hexachlorocyclopentadiene	12.78	237	44958	41.60	ng	97
42) 2,4,6-Trichlorophenol	13.15	196	66405	42.86	ng	94

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF012615\
 Data File : BF077081.D
 Acq On : 27 Jan 2015 9:25
 Operator : TP/IZ
 Sample : G1044-03MSD
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD


Manual Integrations
 APPROVED

apatel
 1/28/2015 5:42:50 PM

Quant Time: Jan 28 14:21:22 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 23 15:21:24 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,5-Trichlorophenol	13.26	196	69810	44.18	ng	# 92
45) 1,1'-Biphenyl	13.55	154	281953	44.22	ng	98
46) 2-Chloronaphthalene	13.52	162	221502	42.21	ng	98
47) 2-Nitroaniline	13.88	65	79653	50.04	ng	86
48) Acenaphthylene	14.47	152	361477	40.84	ng	98
49) Dimethylphthalate	14.43	163	294878	48.34	ng	97
50) 2,6-Dinitrotoluene	14.52	165	57444	47.13	ng	# 88
51) Acenaphthene	14.89	154	216321	43.08	ng	97
52) 3-Nitroaniline	14.87	138	18787	13.17	ng	# 89
53) 2,4-Dinitrophenol	15.15	184	39133	70.12	ng	# 90
54) Dibenzofuran	15.32	168	321256	43.92	ng	99
55) 4-Nitrophenol	15.53	139	33568m	35.11	ng	
56) 2,4-Dinitrotoluene	15.45	165	78805	43.35	ng	93
57) Fluorene	16.06	166	264472	42.67	ng	98
58) 2,3,4,6-Tetrachlorophenol	15.67	232	53654	46.63	ng	97
59) Diethylphthalate	16.06	149	305629	49.57	ng	99
60) 4-Chlorophenyl-phenylether	16.15	204	111712	44.06	ng	97
61) 4-Nitroaniline	16.22	138	49167	32.48	ng	94
62) Azobenzene	16.45	77	290006	45.15	ng	96
64) 4,6-Dinitro-2-methylphenol	16.29	198	29587	34.06	ng	# 66
65) n-Nitrosodiphenylamine	16.41	169	293536	59.06	ng	99
66) 4-Bromophenyl-phenylether	17.01	248	65373	46.29	ng	96
67) Hexachlorobenzene	17.03	284	67528	45.37	ng	91
68) Atrazine	17.42	200	72716	48.21	ng	99
69) Pentachlorophenol	17.41	266	68270	98.27	ng	98
70) Phenanthrene	17.68	178	370173	42.58	ng	98
71) Anthracene	17.75	178	361287	42.62	ng	99
72) Carbazole	18.05	167	354561	43.12	ng	99
73) Di-n-butylphthalate	18.64	149	461259	48.47	ng	98
74) Fluoranthene	19.33	202	392386	44.05	ng	97
77) Pyrene	19.61	202	409401	45.81	ng	99
79) Butylbenzylphthalate	20.55	149	208476	51.51	ng	98
80) Benzo(a)anthracene	21.13	228	357251	43.67	ng	98
81) 3,3'-Dichlorobenzidine	21.16	252	10117	3.89	ng	# 91
82) Chrysene	21.18	228	334771	44.87	ng	98
83) Bis(2-ethylhexyl)phthalate	21.29	149	291192	52.00	ng	99
84) Di-n-octyl phthalate	22.05	149	491865	50.61	ng	# 98
85) Indeno(1,2,3-cd)pyrene	23.89	276	343611	43.46	ng	# 84
87) Benzo(b)fluoranthene	22.38	252	373289	47.37	ng	100
88) Benzo(k)fluoranthene	22.41	252	270109m	39.23	ng	
89) Benzo(a)pyrene	22.73	252	318675	45.84	ng	99
90) Dibenzo(a,h)anthracene	23.91	278	288691	45.26	ng	99
91) Benzo(g,h,i)perylene	24.16	276	286805	45.23	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 7.77 min Scan# 585

Delta R.T. 0.01 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F


ClientSampleId :

SP-1AMSD




Tgt Ion:	152	Resp:	35587
Ion Ratio	Lower	Upper	
152	100		
150	154.3	121.7	182.5
115	62.1	47.0	70.6

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 22.07 ng


RT: 1.16 min Scan# 6

Delta R.T. 0.01 min



Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion:	88	Resp:	20453
Ion Ratio	Lower	Upper	
88	100		
58	104.3	77.0	115.4
43	35.1	25.6	38.4



Abundance Ion 88.00 (87.70 to 88.70): BF0
Ion 58.00 (57.70 to 58.70): BF0
Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 11.93 ng

RT: 1.64 min Scan# 48

Delta R.T. 0.06 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F





ClientSampleId :

SP-1AMSD

Tgt Ion:	79	Resp:	28605
Ion Ratio	Lower	Upper	
79	100		
52	78.3	70.1	105.1
51	38.1	35.1	52.7

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#4

n-Nitrosodimethylamine

Concen: 30.84 ng




RT: 1.58 min Scan# 43

Delta R.T. 0.02 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion:	42	Resp:	36632
Ion Ratio	Lower	Upper	
42	100		
74	93.7	92.4	138.6
44	9.0	6.2	9.2





#5

2-Fluorophenol

Concen: 64.81 ng

RT: 4.82 min Scan# 327

Delta R.T. 0.05 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F


ClientSampleId :

SP-1AMSD

Tgt Ion:	112	Resp:	140280
Ion Ratio		Lower	Upper
112	100		
64	62.9	54.2	81.4
63	35.5	28.4	42.6

Manual Integrations
APPROVED




apatel
1/28/2015 5:42:50 PM



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 43.46 ng


RT: 7.06 min Scan# 523

Delta R.T. 0.02 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25



Tgt Ion:	93	Resp:	164227
Ion Ratio		Lower	Upper
93	100		
66	42.6	35.0	52.4
65	23.2	17.7	26.5




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 43.51 ng

RT: 7.21 min Scan# 536

Delta R.T. 0.05 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F





ClientSampleId :

SP-1AMSD

Tgt Ion:	99	Resp:	118809
Ion Ratio	Lower	Upper	
99	100		
42	18.4	15.0	22.4
71	40.3	30.5	45.7

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#8

2-Chlorophenol

Concen: 45.02 ng




RT: 7.32 min Scan# 545


Delta R.T. 0.02 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion:	128	Resp:	112341
Ion Ratio	Lower	Upper	
128	100		
130	32.6	11.0	51.0
64	54.1	29.3	69.3





#9

Benzaldehyde

Concen: 3.54 ng

RT: 6.79 min Scan# 499

Delta R.T. 0.02 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F





ClientSampleId :

SP-1AMSD

Tgt Ion	Ratio	Lower	Upper
77	100		
105	714.0	81.4	121.4#
106	77.7	72.8	112.8

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#10

Phenol

Concen: 30.10 ng m




RT: 7.25 min Scan# 539


Delta R.T. 0.05 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Lower	Upper
94	100		
65	32.8	11.3	51.3
66	53.6	19.5	59.5





#11





bis(2-Chloroethyl)ether
Concen: 58.53 ng
RT: 7.32 min Scan# 545
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ratio	Lower	Upper
93	100		
63	109.3	111.3	151.3#
95	32.6	13.8	53.8

Manual Integrations
APPROVED




apatel
1/28/2015 5:42:50 PM



#12

1,3-Dichlorobenzene
Concen: 42.24 ng
RT: 7.61 min Scan# 571
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Lower	Upper
146	100		
148	62.9	48.1	72.1
75	33.4	26.8	40.2






#13
 1,4-Dichlorobenzene
 Concen: 52.69 ng
 RT: 7.81 min Scan# 588
 Delta R.T. 0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD




Tgt Ion	Ratio	Resp	Lower	Upper
146	100	158628		
148	65.1	51.5	77.3	
111	43.3	33.9	50.9	

Manual Integrations
 APPROVED

apatel
 1/28/2015 5:42:50 PM




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E






#14
 1,2-Dichlorobenzene
 Concen: 55.31 ng
 RT: 8.13 min Scan# 616
 Delta R.T. 0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
146	100	153418		
148	62.6	51.3	76.9	
111	43.6	34.5	51.7	



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 29.59 ng

RT: 8.24 min Scan# 626

Delta R.T. 0.03 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F





ClientSampled :

SP-1AMSD

Tgt Ion:	79	Resp:	65379
Ion Ratio	Lower	Upper	
79	100		
108	70.5	57.2	85.8
77	61.4	51.2	76.8

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#16

2,2'-oxybis(1-Chloropropane)

Concen: 43.22 ng




RT: 8.56 min Scan# 654


Delta R.T. 0.00 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion:	45	Resp:	131805
Ion Ratio	Lower	Upper	
45	100		
77	17.8	36.1	76.1#
79	13.5	30.5	70.5#






#17
2-Methylphenol
Concen: 38.30 ng
RT: 8.60 min Scan# 657
Delta R.T. 0.03 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD




Tgt Ion:	107	Resp:	78108
Ion Ratio	Lower	Upper	
107	100		
108	114.7	89.7	134.5
77	52.6	50.1	75.1
79	49.6	45.0	67.6

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM






Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF0
Ion 79.00 (78.70 to 79.70): BF0






#18
Hexachloroethane
Concen: 43.33 ng
RT: 8.87 min Scan# 681
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion:	117	Resp:	45374
Ion Ratio	Lower	Upper	
117	100		
119	93.7	77.1	115.7
201	83.9	66.8	100.2

Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 46.02 ng
RT: 8.86 min Scan# 680
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Manual Integrations
APPROVED


apatel
1/28/2015 5:42:50 PM


Tgt Ion	Ratio	Lower	Upper
70	100		
42	55.4	42.1	63.1
101	9.5	7.8	11.8
130	19.6	16.2	24.4



#20
3+4-Methylphenols
Concen: 57.43 ng
RT: 8.97 min Scan# 690
Delta R.T. 0.03 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Lower	Upper
107	100		
108	83.0	63.5	103.5
77	34.9	14.0	54.0
79	26.4	9.9	49.9






#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.69 min Scan# 840
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampled :
SP-1AMSD




Tgt Ion	Ratio	Resp	Lower	Upper
136	100	151663		
137	11.8	8.1	12.1	
54	8.2	7.0	10.4	
68	4.2	3.7	5.5	

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM






Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0




#22
Acetophenone
Concen: 51.17 ng
RT: 8.77 min Scan# 672
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
105	100	190099		
71	0.4	0.0	0.0#	
51	30.3	23.3	34.9	
120	20.5	16.1	24.1	

Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BF0
Ion 51.00 (50.70 to 51.70): BF0
Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 94.93 ng
RT: 9.09 min Scan# 700
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	259879		
128	43.5	33.1	49.7	
54	55.1	48.3	72.5	




Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#24
Nitrobenzene
Concen: 49.04 ng
RT: 9.12 min Scan# 703
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
77	100	136754		
123	42.4	35.2	52.8	
65	15.6	11.4	17.2	





#25

Isophorone

Concen: 46.29 ng

RT: 9.74 min Scan# 757

Delta R.T. 0.02 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F





ClientSampleId :

SP-1AMSD

Tgt Ion:	82	Resp:	230797
Ion Ratio		Lower	Upper
82	100		
95	8.7	6.6	10.0
138	18.5	12.6	18.8

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#26

2-Nitrophenol

Concen: 44.38 ng




RT: 9.85 min Scan# 767


Delta R.T. 0.00 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion:	139	Resp:	63175
Ion Ratio		Lower	Upper
139	100		
109	36.2	21.3	31.9#
65	60.8	39.4	59.0#









#27
 2,4-Dimethylphenol
 Concen: 54.53 ng
 RT: 10.16 min Scan# 794
 Delta R.T. 0.02 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
122	100	117587		
107	121.4	95.2	142.8	
121	58.0	42.9	64.3	




Manual Integrations
 APPROVED


apatel
 1/28/2015 5:42:50 PM



#28
 bis(2-Chloroethoxy)methane
 Concen: 52.53 ng
 RT: 10.36 min Scan# 811
 Delta R.T. 0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	148856		
95	33.8	25.6	38.4	
123	10.9	8.9	13.3	









#29
 2,4-Dichlorophenol
 Concen: 44.39 ng
 RT: 10.48 min Scan# 822
 Delta R.T. 0.02 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
162	100	89215		
164	61.7	43.8	83.8	
98	38.6	22.5	62.5	




Manual Integrations
 APPROVED


apatel
 1/28/2015 5:42:50 PM



#30
 1,2,4-Trichlorobenzene
 Concen: 44.55 ng
 RT: 10.60 min Scan# 832
 Delta R.T. 0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
180	100	99458		
182	90.8	73.6	110.4	
145	33.0	26.6	39.8	









#31
Naphthalene
Concen: 90.48 ng
RT: 10.73 min Scan# 844
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampled :
SP-1AMSD

Tgt Ion	Ratio	Resp Lower	Upper
128	100		
129	11.5	9.0	13.6
127	13.0	10.2	15.2




Manual Integrations
APPROVED


apatel
1/28/2015 5:42:50 PM



#32
Benzoic acid
Concen: 64.07 ng m
RT: 10.61 min Scan# 833
Delta R.T. 0.08 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp Lower	Upper
122	100		
105	74.9	111.3	151.3#
77	70.1	79.2	119.2#






#33
4-Chloroaniline
Concen: 13.50 ng
RT: 10.99 min Scan# 866
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD




Tgt Ion:	127	Resp:	42594
Ion Ratio	Lower	Upper	
127	100		
129	35.0	25.3	37.9
65	40.2	27.8	41.6
92	23.9	16.5	24.7

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0






#34
Hexachlorobutadiene
Concen: 42.87 ng
RT: 11.10 min Scan# 876
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion:	225	Resp:	56786
Ion Ratio	Lower	Upper	
225	100		
223	58.8	49.1	73.7
227	61.4	47.5	71.3



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#36
 4-Chloro-3-methylphenol
 Concen: 102.11 ng
 RT: 12.32 min Scan# 983
 Delta R.T. 0.05 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampled :
 SP-1AMSD

Tgt Ion	Ratio	Resp Lower	Upper
107	100		
144	10.8	20.0	30.0#
142	33.6	63.5	95.3#




Manual Integrations
 APPROVED


apatel
 1/28/2015 5:42:50 PM



#37
 2-Methylnaphthalene
 Concen: 54.15 ng
 RT: 12.39 min Scan# 989
 Delta R.T. 0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp Lower	Upper
142	100		
141	88.7	74.5	111.7
115	37.2	29.7	44.5









#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.81 min Scan# 1201
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
164	100	86014		
162	103.2	82.4	123.6	
160	43.9	34.8	52.2	




Manual Integrations
APPROVED


apatel
1/28/2015 5:42:50 PM



#39
1,2,4,5-Tetrachlorobenzene
Concen: 44.74 ng
RT: 12.79 min Scan# 1024
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
216	100	95135		
214	76.8	64.9	97.3	
179	23.6	18.6	28.0	
108	24.8	18.1	27.1	









#40
Hexachlorocyclopentadiene
Concen: 41.60 ng
RT: 12.78 min Scan# 1023
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
237	100		
235	62.9	40.4	80.4
272	12.3	0.0	33.2




Manual Integrations
APPROVED


apatel
1/28/2015 5:42:50 PM



#41
2,4,6-Tribromophenol
Concen: 144.58 ng
RT: 16.54 min Scan# 1352
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp Lower	Resp Upper
330	100		
332	97.4	77.8	116.6
141	50.9	38.5	57.7






#42
 2,4,6-Trichlorophenol
 Concen: 42.86 ng
 RT: 13.15 min Scan# 1055
 Delta R.T. 0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampled :
 SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	66405		
198	94.9	72.0	108.0	
200	25.7	23.4	35.2	

Manual Integrations
 APPROVED

apatel
 1/28/2015 5:42:50 PM






Abundance

Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E



#43
 2,4,5-Trichlorophenol
 Concen: 44.18 ng
 RT: 13.26 min Scan# 1065
 Delta R.T. 0.05 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	69810		
198	90.5	76.6	115.0	
97	57.9	43.4	65.0	
132	39.9	23.7	35.5	




Abundance


Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 97.00 (96.70 to 97.70): BF0

Ion 132.00 (131.70 to 132.70): E









#44
2-Fluorobiphenyl
Concen: 92.61 ng
RT: 13.34 min Scan# 1072
Delta R.T. 0.00 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion:	172	Resp:	523446
Ion Ratio	Lower	Upper	
172	100		
171	36.2	26.6	40.0
170	23.3	18.0	27.0




Manual Integrations
APPROVED


apatel
1/28/2015 5:42:50 PM



#45
1,1'-Biphenyl
Concen: 44.22 ng
RT: 13.55 min Scan# 1090
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion:	154	Resp:	281953
Ion Ratio	Lower	Upper	
154	100		
153	41.1	20.2	60.2
76	10.5	0.0	29.0









#46
 2-Chloronaphthalene
 Concen: 42.21 ng
 RT: 13.52 min Scan# 1088
 Delta R.T. 0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
162	100	221502		
127	39.7	33.8	50.6	
164	33.5	26.8	40.2	




Manual Integrations
 APPROVED


apatel
 1/28/2015 5:42:50 PM



#47
 2-Nitroaniline
 Concen: 50.04 ng
 RT: 13.88 min Scan# 1119
 Delta R.T. 0.02 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
65	100	79653		
92	66.0	47.3	70.9	
138	105.8	71.4	107.2	









#48
Acenaphthylene
Concen: 40.84 ng
RT: 14.47 min Scan# 1171
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
152	100	361477		
151	19.3	15.9	23.9	
153	11.9	10.2	15.4	




Manual Integrations
APPROVED


apatel
1/28/2015 5:42:50 PM



#49
Dimethylphthalate
Concen: 48.34 ng
RT: 14.43 min Scan# 1167
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
163	100	294878		
194	3.0	2.6	4.0	
164	10.7	7.5	11.3	









#50
2,6-Dinitrotoluene
Concen: 47.13 ng
RT: 14.52 min Scan# 1175
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	57444		
63	73.1	48.1	72.1#	
89	63.8	47.2	70.8	




Manual Integrations
APPROVED


apatel
1/28/2015 5:42:50 PM



#51
Acenaphthene
Concen: 43.08 ng
RT: 14.89 min Scan# 1208
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	216321		
153	112.8	87.0	130.6	
152	53.2	43.5	65.3	





#52

3-Nitroaniline

Concen: 13.17 ng

RT: 14.87 min Scan# 1206

Delta R.T. 0.01 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F


ClientSampleId :

SP-1AMSD

Tgt Ion	Ratio	Lower	Upper
138	100		
108	17.0	8.3	12.5#
92	111.9	80.9	121.3

Manual Integrations
APPROVED




apatel
1/28/2015 5:42:50 PM



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0



#53

2,4-Dinitrophenol

Concen: 70.12 ng

RT: 15.15 min Scan# 1230

Delta R.T. 0.01 min

Lab File: BF077081.D




Acq: 27 Jan 2015 9:25


Tgt Ion	Ratio	Lower	Upper
184	100		
63	95.5	63.0	94.6#
154	81.5	65.5	98.3

Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 154.00 (153.70 to 154.70): E





#54

Dibenzenofuran

Concen: 43.92 ng

RT: 15.32 min Scan# 1245

Delta R.T. 0.01 min

Lab File: BF077081.D


Acq: 27 Jan 2015 9:25

Instrument :

BNA_F

ClientSampled :




SP-1AMSD



Tgt Ion:	168	Resp:	321256
Ion Ratio	Lower	Upper	
168	100		
139	42.3	33.4	50.2
169	12.4	10.6	15.8

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#55

4-Nitrophenol

Concen: 35.11 ng m




RT: 15.53 min Scan# 1264


Delta R.T. 0.08 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion:	139	Resp:	33568
Ion Ratio	Lower	Upper	
139	100		
109	82.0	61.0	101.0
65	112.1	96.6	136.6






#56
2,4-Dinitrotoluene
Concen: 43.35 ng
RT: 15.45 min Scan# 1257
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
165	100		78805		
63	51.1	44.8	67.2		
89	76.2	56.2	84.2		
182	1.7	1.5	2.3		

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0


Ion 89.00 (88.70 to 89.70): BF0

Ion 182.00 (181.70 to 182.70): E



#57
Fluorene
Concen: 42.67 ng
RT: 16.06 min Scan# 1310
Delta R.T. 0.00 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25



Tgt Ion	Ion	Ratio	Resp	Lower	Upper
166	100		264472		
165	95.9	78.5	117.7		
167	12.8	10.4	15.6		



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#58
 2,3,4,6-Tetrachlorophenol
 Concen: 46.63 ng
 RT: 15.67 min Scan# 1276
 Delta R.T. 0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD

Tgt Ion	Ratio	Lower	Upper
232	100		
131	59.8	45.0	67.6
130	3.7	2.6	3.8
166	39.3	30.9	46.3




Manual Integrations
 APPROVED


apatel
 1/28/2015 5:42:50 PM



#59
 Diethylphthalate
 Concen: 49.57 ng
 RT: 16.06 min Scan# 1310
 Delta R.T. 0.00 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Lower	Upper
149	100		
177	18.7	15.0	22.6
150	12.7	9.4	14.0






#60
4-Chlorophenyl-phenylether
Concen: 44.06 ng
RT: 16.15 min Scan# 1318
Delta R.T. 0.00 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
204	100	111712		
206	30.5	24.2	36.2	
141	79.3	66.7	100.1	

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM






Abundance

Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E



#61
4-Nitroaniline
Concen: 32.48 ng
RT: 16.22 min Scan# 1324
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25




Tgt Ion	Ratio	Resp	Lower	Upper
138	100	49167		
92	51.9	26.7	66.7	
108	88.9	64.5	104.5	


Abundance

Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E






#62
Azobenzene
Concen: 45.15 ng
RT: 16.45 min Scan# 1344
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
77	100		
182	26.1	2.0	42.0
105	23.3	2.3	42.3
51	31.2	10.7	50.7

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM






Abundance Ion 77.00 (76.70 to 77.70): BF076858.D (-1354) (-)

Ion 182.00 (181.70 to 182.70): BF076858.D (-1354) (-)

Ion 105.00 (104.70 to 105.70): BF076858.D (-1354) (-)

Ion 51.00 (50.70 to 51.70): BF076858.D (-1354) (-)






#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.64 min Scan# 1448
Delta R.T. 0.00 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25


Tgt Ion	Ratio	Resp Lower	Resp Upper
188	100		
94	9.8	6.0	9.0#
80	9.6	6.9	10.3

Abundance Ion 188.00 (187.70 to 188.70): BF076858.D (-1460) (-)

Ion 94.00 (93.70 to 94.70): BF076858.D (-1460) (-)

Ion 80.00 (79.70 to 80.70): BF076858.D (-1460) (-)









#64
 4,6-Dinitro-2-methylphenol
 Concen: 34.06 ng
 RT: 16.29 min Scan# 1330
 Delta R.T. 0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
198	100		
51	66.0	24.2	64.2#
105	65.7	24.0	64.0#




Manual Integrations
 APPROVED


apatel
 1/28/2015 5:42:50 PM



#65
 n-Nitrosodiphenylamine
 Concen: 59.06 ng
 RT: 16.41 min Scan# 1341
 Delta R.T. 0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp Lower	Resp Upper
169	100		
168	66.8	53.8	80.6
167	36.3	27.8	41.6






#66
4-Bromophenyl-phenylether
Concen: 46.29 ng
RT: 17.01 min Scan# 1393
Delta R.T. 0.00 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD




Tgt Ion	Ratio	Lower	Upper
248	100		
250	97.8	77.1	115.7
141	104.6	88.2	132.4

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM






Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E




#67
Hexachlorobenzene
Concen: 45.37 ng
RT: 17.03 min Scan# 1395
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Lower	Upper
284	100		
142	40.5	38.2	57.4
249	32.0	28.2	42.2

Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E









#68
 Atrazine
 Concen: 48.21 ng
 RT: 17.42 min Scan# 1429
 Delta R.T. 0.03 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
200	100		
173	25.6	6.6	46.6
215	42.5	22.9	62.9




Manual Integrations
 APPROVED


apatel
 1/28/2015 5:42:50 PM



#69
 Pentachlorophenol
 Concen: 98.27 ng
 RT: 17.41 min Scan# 1428
 Delta R.T. 0.02 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp Lower	Resp Upper
266	100		
268	62.1	48.6	72.8
264	60.3	49.5	74.3





#70

Phenanthrene

Concen: 42.58 ng

RT: 17.68 min Scan# 1452

Delta R.T. 0.01 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F





ClientSampleId :

SP-1AMSD

Tgt Ion:	178	Resp:	370173
Ion Ratio	Lower	Upper	
178	100		
176	18.4	15.1	22.7
179	14.9	12.7	19.1

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#71

Anthracene

Concen: 42.62 ng




RT: 17.75 min Scan# 1458


Delta R.T. 0.00 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion:	178	Resp:	361287
Ion Ratio	Lower	Upper	
178	100		
176	19.4	15.4	23.0
179	15.7	11.9	17.9





#72

Carbazole

Concen: 43.12 ng

RT: 18.05 min Scan# 1484

Delta R.T. 0.01 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F





ClientSampleId :

SP-1AMSD

Tgt Ion:	167	Resp:	354561
Ion Ratio	Lower	Upper	
167	100		
166	22.0	17.0	25.6
139	14.3	11.3	16.9

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#73

Di-n-butylphthalate

Concen: 48.47 ng




RT: 18.64 min Scan# 1536


Delta R.T. 0.00 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion:	149	Resp:	461259
Ion Ratio	Lower	Upper	
149	100		
150	9.3	6.9	10.3
104	5.5	3.7	5.5





#74

Fluoranthene

Concen: 44.05 ng

RT: 19.33 min Scan# 1596

Delta R.T. 0.00 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F





ClientSampleId :

SP-1AMSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
202	100		
101	8.6	0.0	30.7
203	17.6	0.0	36.9

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#75

Chrysene-d12

Concen: 20.00 ng




RT: 21.15 min Scan# 1755


Delta R.T. 0.00 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp Lower	Resp Upper
240	100		
120	11.5	10.4	15.6
236	24.8	20.2	30.2





#77

Pyrene

Concen: 45.81 ng

RT: 19.61 min Scan# 1621

Delta R.T. 0.01 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F





ClientSampleId :

SP-1AMSD

Tgt Ion:	202	Resp:	409401
Ion Ratio	Lower	Upper	
202	100		
200	21.0	16.7	25.1
203	18.0	14.0	21.0

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#78

Terphenyl-d14

Concen: 84.61 ng




RT: 19.88 min Scan# 1644


Delta R.T. 0.00 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion:	244	Resp:	479268
Ion Ratio	Lower	Upper	
244	100		
212	8.1	6.3	9.5
122	14.7	10.1	15.1









#79
Butylbenzylphthalate
Concen: 51.51 ng
RT: 20.55 min Scan# 1703
Delta R.T. 0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
91	64.3	50.6	76.0
206	17.3	12.8	19.2




Manual Integrations
APPROVED


apatel
1/28/2015 5:42:50 PM



#80
Benzo(a)anthracene
Concen: 43.67 ng
RT: 21.13 min Scan# 1754
Delta R.T. 0.00 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp Lower	Resp Upper
228	100		
226	26.9	22.1	33.1
229	20.2	15.5	23.3





#81

3,3'-Dichlorobenzidine

Concen: 3.89 ng

RT: 21.16 min Scan# 1756

Delta R.T. 0.01 min

Lab File: BF077081.D


Acq: 27 Jan 2015 9:25

Instrument :

BNA_F

ClientSampleId :

SP-1AMSD



Tgt Ion: 252 Resp: 10117

Ion Ratio Lower Upper



252 100

254 57.2 50.2 75.4

126 22.2 11.5 17.3#

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#82

Chrysene

Concen: 44.87 ng

RT: 21.18 min Scan# 1758

Delta R.T. 0.00 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25




Tgt Ion: 228 Resp: 334771


Ion Ratio Lower Upper

228 100

226 30.0 23.4 35.0

229 19.9 15.4 23.0









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 52.00 ng
 RT: 21.29 min Scan# 1768
 Delta R.T. 0.00 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	291192		
167	26.0	20.6	30.8	
279	2.9	2.2	3.2	




Manual Integrations
 APPROVED


apatel
 1/28/2015 5:42:50 PM



#84
 Di-n-octyl phthalate
 Concen: 50.61 ng
 RT: 22.05 min Scan# 1834
 Delta R.T. -0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	491865		
167	1.8	1.1	1.7#	
43	9.1	6.6	10.0	









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 43.46 ng
 RT: 23.89 min Scan# 1995
 Delta R.T. -0.01 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Instrument :
 BNA_F
 ClientSampleId :
 SP-1AMSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
276	100		
138	26.7	14.6	22.0#
277	24.5	14.8	22.2#




Manual Integrations
 APPROVED


apatel
 1/28/2015 5:42:50 PM



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 22.79 min Scan# 1899
 Delta R.T. 0.00 min
 Lab File: BF077081.D
 Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp Lower	Resp Upper
264	100		
260	24.1	19.0	28.4
265	21.6	18.1	27.1





#87

Benzo(b)fluoranthene

Concen: 47.37 ng

RT: 22.38 min Scan# 1863

Delta R.T. -0.01 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F





ClientSampleId :

SP-1AMSD

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.5	26.3
125	11.8	9.4	14.0

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM



#88

Benzo(k)fluoranthene

Concen: 39.23 ng m




RT: 22.41 min Scan# 1866


Delta R.T. 0.00 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.6	17.2	25.8
125	10.3	7.1	10.7









#89
Benzo(a)pyrene
Concen: 45.84 ng
RT: 22.73 min Scan# 1894
Delta R.T. 0.00 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Instrument :
BNA_F
ClientSampleId :
SP-1AMSD

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	318675		
253	22.0	17.9	26.9	
125	8.3	7.2	10.8	




Manual Integrations
APPROVED




apatel
1/28/2015 5:42:50 PM



#90
Dibenzo(a,h)anthracene
Concen: 45.26 ng
RT: 23.91 min Scan# 1997
Delta R.T. -0.01 min
Lab File: BF077081.D
Acq: 27 Jan 2015 9:25

Tgt Ion	Ratio	Resp	Lower	Upper
278	100	288691		
139	13.2	10.5	15.7	
279	23.6	19.3	28.9	





#91

Benzo(g,h,i)perylene

Concen: 45.23 ng

RT: 24.16 min Scan# 2019

Delta R.T. -0.01 min

Lab File: BF077081.D

Acq: 27 Jan 2015 9:25

Instrument :

BNA_F

ClientSampleId :

SP-1AMSD

Tgt	Ion	Ratio	Resp	Lower	Upper
276	100		286805		
277	24.7	19.5	29.3		
138	23.1	15.0	22.6		

Manual Integrations
APPROVED

apatel
1/28/2015 5:42:50 PM

