


Data Path : Z:\HPCHEM1\BNA F\Data\BF012617\
 Data File : BF092508.D
 Acq On : 26 Jan 2017 7:07
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF012417.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jan 24 13:48:07 2017



AutoFind: Scans 846, 847, 848; Background Corrected with Scan 843

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.3	37602	PASS
68	69	0.00	2	1.5	443	PASS
69	198	0.00	100	38.7	30176	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.2	36016	PASS
197	198	0.00	2	0.8	625	PASS
198	198	100	100	100.0	77876	PASS
199	198	5	9	6.3	4930	PASS
275	198	10	60	22.4	17478	PASS
365	198	1	100	2.2	1707	PASS
441	442	0.01	24	15.7	9885	PASS
442	198	50	100	80.7	62866	PASS
443	442	15	24	17.9	11274	PASS