

Data Path : U:\HPCHEM1\BNA F\DATA\BF012618\
 Data File : BF102473.D
 Acq On : 26 Jan 2018 17:21
 Operator : SJ/JU
 Sample : J1020-15 5X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PL-02-012418-D

Manual Integrations
 APPROVED

Sohil
 1/27/2018 1:46:58 PM

Quant Time: Jan 26 18:36:37 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF012218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 26 16:18:07 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	155952	20.00	ng	0.00
21) Naphthalene-d8	8.00	136	645911	20.00	ng	0.00
38) Acenaphthene-d10	9.76	164	275965	20.00	ng	0.00
63) Phenanthrene-d10	11.25	188	412510	20.00	ng	0.00
75) Chrysene-d12	13.89	240	300420	20.00	ng	0.00
86) Perylene-d12	15.32	264	228974	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	247690	24.08	ng	0.00
7) Phenol-d6	6.36	99	306878	24.75	ng	0.00
23) Nitrobenzene-d5	7.28	82	135527	12.06	ng	0.00
41) 2,4,6-Tribromophenol	10.55	330	37889	13.86	ng	0.00
44) 2-Fluorobiphenyl	9.07	172	311946	16.62	ng	0.00
78) Terphenyl-d14	12.83	244	212242	15.93	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.47	163	54376	2.384	ng	99
70) Phenanthrene	11.27	178	114775	4.775	ng	98
74) Fluoranthene	12.46	202	170049	6.937	ng	98
77) Pyrene	12.69	202	215701	9.286	ng	97
80) Benzo(a)anthracene	13.87	228	83520	4.516	ng	98
82) Chrysene	13.91	228	82980	4.418	ng	99
87) Benzo(b)fluoranthene	14.91	252	70672m	4.762	ng	
89) Benzo(a)pyrene	15.26	252	53190	3.974	ng	# 82
91) Benzo(g,h,i)perylene	17.16	276	28474	2.660	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : U:\HPCHEM1\BNA F\DATA\BF012618\
Data File : BF102473.D
Acq On : 26 Jan 2018 17:21
Operator : SJ/JU
Sample : J1020-15 5X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PL-02-012418-D

Manual Integrations
APPROVED

Sohil
1/27/2018 1:46:58 PM

Quant Time: Jan 26 18:36:37 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF012218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 26 16:18:07 2018
Response via : Initial Calibration

