

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF012718\
 Data File : BF102527.D
 Acq On : 27 Jan 2018 22:57
 Operator : SJ/JU
 Sample : J1290-09
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-59-9.5-10.0

Quant Time: Jan 28 01:19:25 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF012218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 20:48:40 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	159806	20.00	ng	0.00
21) Naphthalene-d8	8.00	136	622891	20.00	ng	0.00
38) Acenaphthene-d10	9.75	164	245738	20.00	ng	0.00
63) Phenanthrene-d10	11.24	188	347930	20.00	ng	0.00
75) Chrysene-d12	13.88	240	285713	20.00	ng	0.00
86) Perylene-d12	15.31	264	222089	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	1017222	96.52	ng	0.01
7) Phenol-d6	6.36	99	1233549	97.08	ng	0.00
23) Nitrobenzene-d5	7.29	82	738231	68.11	ng	0.00
41) 2,4,6-Tribromophenol	10.55	330	178051	73.12	ng	0.00
44) 2-Fluorobiphenyl	9.07	172	1233637	73.81	ng	0.00
78) Terphenyl-d14	12.82	244	796695	62.86	ng	0.00
Target Compounds						
10) Phenol	6.37	94	48853	3.522	ng	98
49) Dimethylphthalate	9.47	163	127482	6.276	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed