

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF012718\
 Data File : BF102551.D
 Acq On : 28 Jan 2018 10:37
 Operator : SJ/JU
 Sample : J1321-05
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OWL-C6-GW

Quant Time: Jan 29 01:00:20 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF012218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 20:48:40 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.71	152	169879	20.00	ng	0.00
21) Naphthalene-d8	7.99	136	650783	20.00	ng	0.00
38) Acenaphthene-d10	9.75	164	253160	20.00	ng	0.00
63) Phenanthrene-d10	11.23	188	349111	20.00	ng	0.00
75) Chrysene-d12	13.87	240	309591	20.00	ng	0.00
86) Perylene-d12	15.30	264	267540	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	615569	54.94	ng	0.00
7) Phenol-d6	6.35	99	463628	34.32	ng	-0.01
23) Nitrobenzene-d5	7.28	82	935737	82.63	ng	0.00
41) 2,4,6-Tribromophenol	10.54	330	222011	88.51	ng	0.00
44) 2-Fluorobiphenyl	9.07	172	1487456	86.39	ng	0.00
78) Terphenyl-d14	12.82	244	981763	71.49	ng	0.00
Target Compounds						
10) Phenol	6.37	94	32569	2.209	ng	93
49) Dimethylphthalate	9.46	163	96378	4.606	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed