

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012720\
 Data File : BF118723.D
 Acq On : 27 Jan 2020 17:09
 Operator : CG/JU
 Sample : L1008-15 2X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CL-02-012420

Manual Integrations
 APPROVED

mohammad
 1/28/2020 1:13:16 PM

Quant Time: Jan 27 17:55:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 27 14:24:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	307915	20.00	ng	0.00
21) Naphthalene-d8	8.15	136	1093595	20.00	ng	0.00
39) Acenaphthene-d10	9.90	164	557756	20.00	ng	0.00
64) Phenanthrene-d10	11.38	188	883557	20.00	ng	0.00
76) Chrysene-d12	14.02	240	619245	20.00	ng	0.00
87) Perylene-d12	15.49	264	753243	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	626591	37.79	ng	0.00
7) Phenol-d6	6.48	99	750788	34.84	ng	-0.01
23) Nitrobenzene-d5	7.42	82	443268	22.67	ng	0.00
42) 2,4,6-Tribromophenol	10.68	330	199505	30.93	ng	0.00
45) 2-Fluorobiphenyl	9.22	172	890452	26.24	ng	0.00
79) Terphenyl-d14	12.96	244	799754	28.17	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.60	163	81802	2.251	ng	99
71) Phenanthrene	11.40	178	177770	4.170	ng	97
75) Fluoranthene	12.59	202	167337	3.722	ng	97
78) Pyrene	12.82	202	167125	3.975	ng	97
83) Chrysene	14.04	228	73124m	2.048	ng	
88) Benzo(b)fluoranthene	15.06	252	97925m	2.068	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012720\
Data File : BF118723.D
Acq On : 27 Jan 2020 17:09
Operator : CG/JU
Sample : L1008-15 2X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CL-02-012420

Manual Integrations
APPROVED

mohammad
1/28/2020 1:13:16 PM

Quant Time: Jan 27 17:55:22 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 27 14:24:10 2020
Response via : Initial Calibration

