




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012725\
Data File : BF141265.D
Acq On : 27 Jan 2025 11:32
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 27 12:26:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141265.D\data.ms

(70) Pentachlorophenol (C)




11.133min (-0.012) 23981.38 ng

response 229106

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	62.71
264.00	62.70	63.18
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012725\
Data File : BF141265.D
Acq On : 27 Jan 2025 11:32
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 27 12:26:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141265.D\data.ms

(77) Benzidine**12.668min (-0.012) 58178.64 ng****response 885766**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.01
183.00	11.50	12.12
0.00	0.00	0.00