



Data Path : U:\HPCHEM1\BNA F\DATA\BF012818\
Data File : BF102581.D
Acq On : 29 Jan 2018 2:53
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 29 07:34:13 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF012218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 27 20:48:40 2018
Response via : Initial Calibration



TIC: BF102581.D

(76) Benzidine

12.580min (+0.006) 0.00ng

response 389782

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	14.66
183.00	11.60	12.35
0.00	0.00	0.00