

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012818\
 Data File : BF102580.D
 Acq On : 29 Jan 2018 1:32
 Operator : SJ/JU
 Sample : J1010-08RE
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BU-01-012518RE

Quant Time: Jan 29 06:19:38 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF012218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 20:48:40 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	157694	20.00	ng	0.00
21) Naphthalene-d8	8.00	136	581974	20.00	ng	0.00
38) Acenaphthene-d10	9.75	164	204954	20.00	ng	0.00
63) Phenanthrene-d10	11.24	188	327530	20.00	ng	0.00
75) Chrysene-d12	13.88	240	336116	20.00	ng	0.00
86) Perylene-d12	15.31	264	240218	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	1139034	109.52	ng	0.01
7) Phenol-d6	6.37	99	1306165	104.17	ng	0.00
23) Nitrobenzene-d5	7.29	82	774544	76.48	ng	0.00
41) 2,4,6-Tribromophenol	10.55	330	175984	86.66	ng	0.00
44) 2-Fluorobiphenyl	9.07	172	1053129	75.55	ng	0.00
78) Terphenyl-d14	12.82	244	890021	59.69	ng	0.00
Target Compounds						
10) Phenol	6.37	94	56087	4.097	ng	82
49) Dimethylphthalate	9.47	163	108813	6.423	ng	100
74) Fluoranthene	12.45	202	42723	2.195	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed