

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012821\
 Data File : BF123054.D
 Acq On : 28 Jan 2021 11:28
 Operator : JU/CG
 Sample : PB134359BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB134359BS

Manual Integrations
 APPROVED

mohammad
 1/29/2021 10:33:26 AM

Quant Time: Jan 28 14:54:18 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 28 00:48:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	298970	20.00	ng	0.00
21) Naphthalene-d8	8.181	136	1152052	20.00	ng	0.00
39) Acenaphthene-d10	9.939	164	620842	20.00	ng	0.00
64) Phenanthrene-d10	11.433	188	1108888	20.00	ng	0.00
76) Chrysene-d12	14.086	240	893039	20.00	ng	0.00
86) Perylene-d12	15.568	264	815560	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	2160116	111.45	ng	0.02
7) Phenol-d6	6.528	99	2850596	108.51	ng	0.01
23) Nitrobenzene-d5	7.457	82	1632496	71.31	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	815316	120.97	ng	0.00
45) 2-Fluorobiphenyl	9.257	172	3001069	71.86	ng	0.00
79) Terphenyl-d14	13.021	244	3262073	71.77	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.793	88	390976	40.54	ng	95
3) Pyridine	3.534	79	1092326	42.36	ng	95
4) n-Nitrosodimethylamine	3.493	42	471181	43.42	ng	96
6) Aniline	6.557	93	1019072	31.52	ng	99
8) 2-Chlorophenol	6.681	128	920063	43.79	ng	96
9) Benzaldehyde	6.439	77	42222	3.52	ng	98
10) Phenol	6.539	94	1181439	45.35	ng	88
11) bis(2-Chloroethyl)ether	6.634	93	925672	42.69	ng	99
12) 1,3-Dichlorobenzene	6.839	146	1008356	41.16	ng	99
13) 1,4-Dichlorobenzene	6.910	146	1019554	39.96	ng	97
14) 1,2-Dichlorobenzene	7.069	146	946497	39.65	ng	100
15) Benzyl Alcohol	7.034	79	812291	44.10	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.169	45	1366145	40.86	ng	97
17) 2-Methylphenol	7.145	107	801842	44.86	ng	98
18) Hexachloroethane	7.410	117	378239	42.32	ng	95
19) n-Nitroso-di-n-propyla...	7.310	70	659421	41.74	ng	96
20) 3+4-Methylphenols	7.298	107	925951	43.98	ng	95
22) Acetophenone	7.304	105	1232034	41.04	ng	98
24) Nitrobenzene	7.475	77	1019936	43.24	ng	98
25) Isophorone	7.716	82	1878626	44.80	ng	100
26) 2-Nitrophenol	7.792	139	496911	50.95	ng	96
27) 2,4-Dimethylphenol	7.828	122	920564	52.31	ng	98
28) bis(2-Chloroethoxy)met...	7.928	93	1190083	41.60	ng	99
29) 2,4-Dichlorophenol	8.033	162	817957	44.45	ng	97
30) 1,2,4-Trichlorobenzene	8.122	180	859848	41.47	ng	99
31) Naphthalene	8.198	128	2671436	42.27	ng	99
32) Benzoic acid	7.957	122	654243	46.34	ng	98
33) 4-Chloroaniline	8.245	127	737068	26.28	ng	99
34) Hexachlorobutadiene	8.322	225	483818	40.40	ng	99
35) Caprolactam	8.622	113	267357m	42.99	ng	
36) 4-Chloro-3-methylphenol	8.728	107	865671	45.24	ng	97
37) 2-Methylnaphthalene	8.892	142	1814671	43.26	ng	100
38) 1-Methylnaphthalene	8.992	142	1700185	42.80	ng	100
40) 1,2,4,5-Tetrachloroben...	9.063	216	787074	40.75	ng	99
41) Hexachlorocyclopentadiene	9.051	237	913984	99.69	ng	98
43) 2,4,6-Trichlorophenol	9.169	196	594277	44.88	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012821\
 Data File : BF123054.D
 Acq On : 28 Jan 2021 11:28
 Operator : JU/CG
 Sample : PB134359BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB134359BS

Manual Integrations
 APPROVED

mohammad
 1/29/2021 10:33:26 AM

Quant Time: Jan 28 14:54:18 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 28 00:48:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.216	196	611355	44.15	ng	99
46) 1,1'-Biphenyl	9.363	154	2148406	42.19	ng	100
47) 2-Chloronaphthalene	9.386	162	1709167	40.09	ng	98
48) 2-Nitroaniline	9.475	65	555499	42.99	ng	85
49) Acenaphthylene	9.804	152	2608855	41.81	ng	99
50) Dimethylphthalate	9.663	163	2030793	41.69	ng	100
51) 2,6-Dinitrotoluene	9.722	165	471980	45.49	ng	95
52) Acenaphthene	9.975	154	1822656	45.48	ng	99
53) 3-Nitroaniline	9.886	138	350719	28.58	ng	89
54) 2,4-Dinitrophenol	9.998	184	443002	92.22	ng	92
55) Dibenzofuran	10.151	168	2361492	40.49	ng	98
56) 4-Nitrophenol	10.051	139	819400	92.39	ng	89
57) 2,4-Dinitrotoluene	10.127	165	625265	46.34	ng	98
58) Fluorene	10.492	166	1784052	38.29	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.263	232	510918	43.51	ng	94
60) Diethylphthalate	10.363	149	1986513	41.48	ng	99
61) 4-Chlorophenyl-phenyle...	10.480	204	858106	39.20	ng	99
62) 4-Nitroaniline	10.510	138	468124	37.97	ng	95
63) Azobenzene	10.645	77	1968469	42.00	ng	97
65) 4,6-Dinitro-2-methylph...	10.533	198	314525	51.04	ng	87
66) n-Nitrosodiphenylamine	10.604	169	1663301	41.61	ng	99
67) 4-Bromophenyl-phenylether	10.974	248	581306	41.16	ng	97
68) Hexachlorobenzene	11.045	284	617906	40.79	ng	95
69) Atrazine	11.127	200	446761	40.48	ng	98
70) Pentachlorophenol	11.233	266	750925	91.47	ng	99
71) Phenanthrene	11.457	178	2663726	38.68	ng	99
72) Anthracene	11.510	178	2751581	41.65	ng	98
73) Carbazole	11.663	167	2471900	40.32	ng	99
74) Di-n-butylphthalate	11.992	149	2960967	40.73	ng	99
75) Fluoranthene	12.651	202	2771716	40.27	ng	98
77) Benzidine	12.768	184	1076747	53.28	ng	99
78) Pyrene	12.880	202	2797694	41.50	ng	98
80) Butylbenzylphthalate	13.498	149	1307108	45.06	ng	96
81) Benzo(a)anthracene	14.074	228	2488677	41.18	ng	100
82) 3,3'-Dichlorobenzidine	14.033	252	783685	41.20	ng	# 98
83) Chrysene	14.115	228	2449739	40.50	ng	97
84) Bis(2-ethylhexyl)phtha...	14.062	149	1675232	43.46	ng	98
85) Di-n-octyl phthalate	14.680	149	2998143	46.31	ng	98
87) Indeno(1,2,3-cd)pyrene	17.074	276	2421451	44.59	ng	96
88) Benzo(b)fluoranthene	15.139	252	2663031	45.24	ng	100
89) Benzo(k)fluoranthene	15.168	252	2282705	41.73	ng	99
90) Benzo(a)pyrene	15.509	252	2223047	43.00	ng	98
91) Dibenzo(a,h)anthracene	17.098	278	2006730	44.89	ng	98
92) Benzo(g,h,i)perylene	17.533	276	1952765	45.08	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012821\
Data File : BF123054.D
Acq On : 28 Jan 2021 11:28
Operator : JU/CG
Sample : PB134359BS
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
PB134359BS

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM


Quant Time: Jan 28 14:54:18 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 28 00:48:07 2021
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 6.892 min Scan# 817
 Delta R.T. 0.000 min
 Lab File: BF123054.D
 Acq: 28 Jan 2021 11:28




Instrument :
 BNA_F
 ClientSampleID :
 PB134359BS




Tgt Ion:152 Resp: 298970
 Ion Ratio Lower Upper
 152 100
 150 155.1 125.8 188.6
 115 62.5 49.3 73.9

Manual Integrations
 APPROVED



mohammad
 1/29/2021 10:33:26 AM




#2
 1,4-Dioxane
 Concen: 40.54 ng
 RT: 2.793 min Scan# 120
 Delta R.T. 0.129 min
 Lab File: BF123054.D
 Acq: 28 Jan 2021 11:28



Tgt Ion: 88 Resp: 390976
 Ion Ratio Lower Upper
 88 100
 58 75.3 56.2 84.2
 43 28.1 21.8 32.6









#3
Pyridine
Concen: 42.36 ng
RT: 3.534 min Scan# 246
Delta R.T. 0.112 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion: 79 Resp: 1092326
Ion Ratio Lower Upper
79 100
52 70.6 53.0 79.4
51 34.8 26.6 39.8

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#4
n-Nitrosodimethylamine
Concen: 43.42 ng
RT: 3.493 min Scan# 239
Delta R.T. 0.112 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion: 42 Resp: 471181
Ion Ratio Lower Upper
42 100
74 139.2 115.2 172.8
44 6.5 5.4 8.0









#5
2-Fluorophenol
Concen: 111.45 ng
RT: 5.528 min Scan# 585
Delta R.T. 0.023 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:112 Resp: 2160116
Ion Ratio Lower Upper
112 100
64 58.0 46.9 70.3
63 29.5 24.5 36.7

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#6
Aniline
Concen: 31.52 ng
RT: 6.557 min Scan# 760
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion: 93 Resp: 1019072
Ion Ratio Lower Upper
93 100
66 38.6 31.3 46.9
65 20.3 15.6 23.4









#7
Phenol-d6
Concen: 108.51 ng
RT: 6.528 min Scan# 755
Delta R.T. 0.012 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion: 99 Resp: 2850596
Ion Ratio Lower Upper
99 100
42 18.1 13.8 20.6
71 34.2 26.9 40.3

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#8
2-Chlorophenol
Concen: 43.79 ng
RT: 6.681 min Scan# 781
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion: 128 Resp: 920063
Ion Ratio Lower Upper
128 100
130 33.0 12.3 52.3
64 49.9 25.5 65.5









#9
Benzaldehyde
Concen: 3.52 ng
RT: 6.439 min Scan# 740
Delta R.T. -0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion: 77 Resp: 42222
Ion Ratio Lower Upper
77 100
105 99.0 80.6 120.6
106 95.3 77.2 117.2

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM




#10
Phenol
Concen: 45.35 ng
RT: 6.539 min Scan# 757
Delta R.T. 0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion: 94 Resp: 1181439
Ion Ratio Lower Upper
94 100
65 27.4 11.2 51.2
66 35.5 25.8 65.8









#11
bis(2-Chloroethyl)ether
Concen: 42.69 ng
RT: 6.634 min Scan# 773
Delta R.T. 0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28





Tgt Ion: 93 Resp: 925672
Ion Ratio Lower Upper
93 100
63 72.7 52.6 92.6
95 33.4 12.6 52.6



#12
1,3-Dichlorobenzene
Concen: 41.16 ng
RT: 6.839 min Scan# 808
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion: 146 Resp: 1008356
Ion Ratio Lower Upper
146 100
148 65.7 51.8 77.8
75 28.0 22.5 33.7



Instrument :
BNA_F
ClientSampleId :
PB134359BS

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM







#13
1,4-Dichlorobenzene
Concen: 39.96 ng
RT: 6.910 min Scan# 820
Delta R.T. -0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:146 Resp: 1019554
Ion Ratio Lower Upper
146 100
148 63.5 52.0 78.0
111 43.3 32.3 48.5

Instrument :
BNA_F
ClientSampleID :
PB134359BS




Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM




#14
1,2-Dichlorobenzene
Concen: 39.65 ng
RT: 7.069 min Scan# 847
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:146 Resp: 946497
Ion Ratio Lower Upper
146 100
148 64.4 51.5 77.3
111 42.1 34.2 51.2









#15
Benzyl Alcohol
Concen: 44.10 ng
RT: 7.034 min Scan# 841
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28





Tgt Ion: 79 Resp: 812291
Ion Ratio Lower Upper
79 100
108 79.0 62.4 93.6
77 62.7 49.0 73.6



#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.86 ng
RT: 7.169 min Scan# 864
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion: 45 Resp: 1366145
Ion Ratio Lower Upper
45 100
77 14.4 0.0 35.6
79 11.2 0.0 32.4



Instrument :
BNA_F
ClientSampleId :
PB134359BS

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM







#17
2-Methylphenol
Concen: 44.86 ng
RT: 7.145 min Scan# 860
Delta R.T. 0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:	107	Resp:	801842
Ion Ratio	Lower	Upper	
107	100		
108	112.7	91.1	136.7
77	43.9	36.3	54.5
79	44.4	36.8	55.2

Instrument :
BNA_F
ClientSampled :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#18
Hexachloroethane
Concen: 42.32 ng
RT: 7.410 min Scan# 905
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:	117	Resp:	378239
Ion Ratio	Lower	Upper	
117	100		
119	94.5	78.4	117.6
201	92.2	79.6	119.4









#19
n-Nitroso-di-n-propylamine
Concen: 41.74 ng
RT: 7.310 min Scan# 888
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:	70	Resp:	659421
Ion	Ratio	Lower	Upper
70	100		
42	57.3	43.0	64.4
101	9.6	7.9	11.9
130	21.5	18.5	27.7

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#20
3+4-Methylphenols
Concen: 43.98 ng
RT: 7.298 min Scan# 886
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28


Tgt Ion:	107	Resp:	925951
Ion	Ratio	Lower	Upper
107	100		
108	92.3	68.7	108.7
77	102.0	89.8	129.8
79	30.5	9.8	49.8





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.181 min Scan# 1036
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Instrument :
BNA_F
ClientSampleId :
PB134359BS






Tgt Ion:136 Resp: 1152052


Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	8.0	5.5	8.3
68	6.1	4.6	6.8

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM






#22
Acetophenone
Concen: 41.04 ng
RT: 7.304 min Scan# 887
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28



Tgt Ion:105 Resp: 1232034


Ion	Ratio	Lower	Upper
105	100		
71	6.7	6.5	9.7
51	29.2	24.3	36.5
120	22.5	18.1	27.1





#23
Nitrobenzene-d5
Concen: 71.31 ng
RT: 7.457 min Scan# 913
Delta R.T. -0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Instrument :
BNA_F
ClientSampleId :
PB134359BS






Tgt Ion: 82 Resp: 1632496


Ion	Ratio	Lower	Upper
82	100		
128	45.7	37.0	55.4
54	49.5	38.4	57.6

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM






#24
Nitrobenzene
Concen: 43.24 ng
RT: 7.475 min Scan# 916
Delta R.T. -0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28



Tgt Ion: 77 Resp: 1019936

Ion	Ratio	Lower	Upper
77	100		
123	47.6	39.3	58.9
65	13.4	10.6	15.8





#25
Isophorone
Concen: 44.80 ng
RT: 7.716 min Scan# 957
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28





Tgt Ion: 82 Resp: 1878626

Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.8	8.6
138	18.2	14.6	21.8

Instrument :
BNA_F
ClientSampleId :
PB134359BS

Manual Integrations
APPROVED




mohammad
1/29/2021 10:33:26 AM




#26
2-Nitrophenol
Concen: 50.95 ng
RT: 7.792 min Scan# 970
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion: 139 Resp: 496911

Ion	Ratio	Lower	Upper
139	100		
109	27.9	24.0	36.0
65	48.1	40.9	61.3





#27

2,4-Dimethylphenol

Concen: 52.31 ng

RT: 7.828 min Scan# 976

Delta R.T. -0.000 min

Lab File: BF123054.D

Acq: 28 Jan 2021 11:28

Tgt Ion:	122	Resp:	920564
Ion	Ratio	Lower	Upper
122	100		
107	112.6	91.8	137.8
121	58.5	47.3	70.9

Instrument :





BNA_F

ClientSampleId :

PB134359BS

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM



#28

bis(2-Chloroethoxy)methane

Concen: 41.60 ng




RT: 7.928 min Scan# 993


Delta R.T. -0.000 min

Lab File: BF123054.D

Acq: 28 Jan 2021 11:28

Tgt Ion:	93	Resp:	1190083
Ion	Ratio	Lower	Upper
93	100		
95	33.2	26.2	39.2
123	12.7	9.2	13.8









#29
2,4-Dichlorophenol
Concen: 44.45 ng
RT: 8.033 min Scan# 1011
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:162 Resp: 817957
Ion Ratio Lower Upper
162 100
164 64.9 43.2 83.2
98 38.1 15.2 55.2

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#30
1,2,4-Trichlorobenzene
Concen: 41.47 ng
RT: 8.122 min Scan# 1026
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:180 Resp: 859848
Ion Ratio Lower Upper
180 100
182 95.2 75.7 113.5
145 29.7 25.0 37.4









#31
Naphthalene
Concen: 42.27 ng
RT: 8.198 min Scan# 1039
Delta R.T. -0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:128 Resp: 2671436
Ion Ratio Lower Upper
128 100
129 12.1 9.0 13.6
127 13.8 10.7 16.1

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#32
Benzoic acid
Concen: 46.34 ng
RT: 7.957 min Scan# 998
Delta R.T. 0.012 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:122 Resp: 654243
Ion Ratio Lower Upper
122 100
105 121.0 104.8 144.8
77 90.2 71.0 111.0









#33
4-Chloroaniline
Concen: 26.28 ng
RT: 8.245 min Scan# 1047
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.2	39.2
65	29.2	22.2	33.2
92	19.0	15.3	22.9

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#34
Hexachlorobutadiene
Concen: 40.40 ng
RT: 8.322 min Scan# 1060
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.7	50.7	76.1
227	63.2	51.7	77.5









#35
Caprolactam
Concen: 42.99 ng m
RT: 8.622 min Scan# 1111
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:113 Resp: 267357
Ion Ratio Lower Upper
113 100
55 172.6 133.7 173.7
56 124.9 95.3 135.3

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#36
4-Chloro-3-methylphenol
Concen: 45.24 ng
RT: 8.728 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:107 Resp: 865671
Ion Ratio Lower Upper
107 100
144 25.3 20.2 30.2
142 77.3 64.4 96.6









#37
2-Methylnaphthalene
Concen: 43.26 ng
RT: 8.892 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:142 Resp: 1814671
Ion Ratio Lower Upper
142 100
141 88.5 70.6 106.0

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#38
1-Methylnaphthalene
Concen: 42.80 ng
RT: 8.992 min Scan# 1174
Delta R.T. -0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28


Tgt Ion:142 Resp: 1700185
Ion Ratio Lower Upper
142 100
141 92.8 74.5 111.7





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.939 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Instrument :
BNA_F
ClientSampleID :
PB134359BS




Tgt Ion:164 Resp: 620842
Ion Ratio Lower Upper
164 100
162 101.6 80.8 121.2
160 45.0 35.4 53.0

Manual Integrations
APPROVED



mohammad
1/29/2021 10:33:26 AM



#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.75 ng
RT: 9.063 min Scan# 1186
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28



Tgt Ion:216 Resp: 787074
Ion Ratio Lower Upper
216 100
214 79.3 63.1 94.7
179 21.7 17.1 25.7
108 21.3 14.6 21.8









#41
Hexachlorocyclopentadiene
Concen: 99.69 ng
RT: 9.051 min Scan# 1184
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:237 Resp: 913984
Ion Ratio Lower Upper
237 100
235 63.0 40.9 80.9
272 13.3 0.0 33.4

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM




#42
2,4,6-Tribromophenol
Concen: 120.97 ng
RT: 10.733 min Scan# 1470
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:330 Resp: 815316
Ion Ratio Lower Upper
330 100
332 95.8 77.5 116.3
141 31.1 23.8 35.8









#43
2,4,6-Trichlorophenol
Concen: 44.88 ng
RT: 9.169 min Scan# 1204
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28





Tgt Ion:196 Resp: 594277
Ion Ratio Lower Upper
196 100
198 96.6 75.1 112.7
200 31.2 23.7 35.5



#44
2,4,5-Trichlorophenol
Concen: 44.15 ng
RT: 9.216 min Scan# 1212
Delta R.T. 0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion:196 Resp: 611355
Ion Ratio Lower Upper
196 100
198 95.9 76.5 114.7
97 51.2 40.1 60.1
132 28.8 23.3 34.9



Instrument :
BNA_F
ClientSampleId :
PB134359BS

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM







#45
2-Fluorobiphenyl
Concen: 71.86 ng
RT: 9.257 min Scan# 1219
Delta R.T. -0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:172 Resp: 3001069
Ion Ratio Lower Upper
172 100
171 36.8 29.6 44.4
170 24.5 19.5 29.3

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#46
1,1'-Biphenyl
Concen: 42.19 ng
RT: 9.363 min Scan# 1237
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:154 Resp: 2148406
Ion Ratio Lower Upper
154 100
153 41.5 21.8 61.8
76 13.2 0.0 33.4









#47
2-Chloronaphthalene
Concen: 40.09 ng
RT: 9.386 min Scan# 1241
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:162 Resp: 1709167
Ion Ratio Lower Upper
162 100
127 41.1 33.8 50.8
164 34.3 26.5 39.7

Instrument :
BNA_F
ClientSampled :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#48
2-Nitroaniline
Concen: 42.99 ng
RT: 9.475 min Scan# 1256
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28


Tgt Ion: 65 Resp: 555499
Ion Ratio Lower Upper
65 100
92 67.7 60.0 90.0
138 103.4 100.6 151.0





#49
Acenaphthylene
Concen: 41.81 ng
RT: 9.804 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Instrument :
BNA_F
ClientSampleId :
PB134359BS




Tgt Ion:152 Resp: 2608855
Ion Ratio Lower Upper
152 100
151 21.3 16.6 25.0
153 13.9 10.9 16.3

Manual Integrations
APPROVED



mohammad
1/29/2021 10:33:26 AM




#50
Dimethylphthalate
Concen: 41.69 ng
RT: 9.663 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28



Tgt Ion:163 Resp: 2030793
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.3 8.1 12.1









#51
2,6-Dinitrotoluene
Concen: 45.49 ng
RT: 9.722 min Scan# 1298
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:165 Resp: 471980
Ion Ratio Lower Upper
165 100
63 50.2 36.6 55.0
89 51.8 39.4 59.0

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM




#52
Acenaphthene
Concen: 45.48 ng
RT: 9.975 min Scan# 1341
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:154 Resp: 1822656
Ion Ratio Lower Upper
154 100
153 109.3 89.0 133.6
152 53.1 42.7 64.1









#53
3-Nitroaniline
Concen: 28.58 ng
RT: 9.886 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28





Tgt Ion:138 Resp: 350719
Ion Ratio Lower Upper
138 100
108 11.5 8.2 12.2
92 118.9 85.4 128.2



#54
2,4-Dinitrophenol
Concen: 92.22 ng
RT: 9.998 min Scan# 1345
Delta R.T. 0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion:184 Resp: 443002
Ion Ratio Lower Upper
184 100
63 55.3 51.4 77.2
154 59.1 49.7 74.5



Instrument :
BNA_F
ClientSampleId :
PB134359BS


Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM



#55
Dibenzofuran
Concen: 40.49 ng
RT: 10.151 min Scan# 1371
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Instrument :
BNA_F
ClientSampled :
PB134359BS




Tgt Ion:168 Resp: 2361492
Ion Ratio Lower Upper
168 100
139 38.0 31.3 46.9
169 13.7 11.2 16.8

Manual Integrations
APPROVED



mohammad
1/29/2021 10:33:26 AM




#56
4-Nitrophenol
Concen: 92.39 ng
RT: 10.051 min Scan# 1354
Delta R.T. 0.012 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion:139 Resp: 819400
Ion Ratio Lower Upper
139 100
109 68.5 45.1 85.1
65 94.0 59.2 99.2





#57
2,4-Dinitrotoluene
Concen: 46.34 ng
RT: 10.127 min Scan# 1367
Delta R.T. 0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Instrument :
BNA_F
ClientSampleId :
PB134359BS




Tgt Ion:165 Resp: 625265
Ion Ratio Lower Upper
165 100
63 41.1 34.3 51.5
89 68.3 55.2 82.8
182 2.7 2.2 3.2

Manual Integrations
APPROVED



mohammad
1/29/2021 10:33:26 AM



#58
Fluorene
Concen: 38.29 ng
RT: 10.492 min Scan# 1429
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion:166 Resp: 1784052
Ion Ratio Lower Upper
166 100
165 99.3 79.4 119.0
167 14.4 10.9 16.3









#59
2,3,4,6-Tetrachlorophenol
Concen: 43.51 ng
RT: 10.263 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28





Tgt Ion:232 Resp: 510918
Ion Ratio Lower Upper
232 100
131 45.5 33.0 49.4
130 2.3 1.7 2.5
166 31.5 23.0 34.4



#60
Diethylphthalate
Concen: 41.48 ng
RT: 10.363 min Scan# 1407
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion:149 Resp: 1986513
Ion Ratio Lower Upper
149 100
177 22.8 18.2 27.4
150 13.2 10.1 15.1




Instrument :
BNA_F
ClientSampleId :
PB134359BS

Manual Integrations
APPROVED




mohammad
1/29/2021 10:33:26 AM




#61
4-Chlorophenyl-phenylether
Concen: 39.20 ng
RT: 10.480 min Scan# 1427
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28





Tgt Ion:204 Resp: 858106
Ion Ratio Lower Upper
204 100
206 34.1 27.0 40.4
141 63.7 50.4 75.6



#62
4-Nitroaniline
Concen: 37.97 ng
RT: 10.510 min Scan# 1432
Delta R.T. 0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion:138 Resp: 468124
Ion Ratio Lower Upper
138 100
92 49.4 31.0 71.0
108 69.2 54.8 94.8




Instrument :
BNA_F
ClientSampleId :
PB134359BS

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM





#63
Azobenzene
Concen: 42.00 ng
RT: 10.645 min Scan# 1455
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28



Tgt Ion: 77 Resp: 1968469


Ion	Ratio	Lower	Upper
77	100		
182	27.4	5.6	45.6
105	23.2	1.8	41.8
51	29.2	8.3	48.3




Instrument :
BNA_F
ClientSampleId :
PB134359BS

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM






#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.433 min Scan# 1589
Delta R.T. 0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28



Tgt Ion: 188 Resp: 1108888

Ion	Ratio	Lower	Upper
188	100		
94	10.3	7.4	11.2
80	11.1	8.2	12.4









#65
4,6-Dinitro-2-methylphenol
Concen: 51.04 ng
RT: 10.533 min Scan# 1436
Delta R.T. 0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:198 Resp: 314525
Ion Ratio Lower Upper
198 100
51 50.6 20.3 60.3
105 48.9 23.0 63.0

Instrument :
BNA_F
ClientSampled :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#66
n-Nitrosodiphenylamine
Concen: 41.61 ng
RT: 10.604 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28


Tgt Ion:169 Resp: 1663301
Ion Ratio Lower Upper
169 100
168 67.6 54.2 81.2
167 37.5 29.2 43.8





#67
4-Bromophenyl-phenylether
Concen: 41.16 ng
RT: 10.974 min Scan# 1511
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Instrument :
BNA_F
ClientSampleId :
PB134359BS




Tgt Ion:248 Resp: 581306
Ion Ratio Lower Upper
248 100
250 95.9 76.8 115.2
141 69.0 59.8 89.8

Manual Integrations
APPROVED



mohammad
1/29/2021 10:33:26 AM




#68
Hexachlorobenzene
Concen: 40.79 ng
RT: 11.045 min Scan# 1523
Delta R.T. 0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28



Tgt Ion:284 Resp: 617906
Ion Ratio Lower Upper
284 100
142 28.0 25.6 38.4
249 30.4 25.1 37.7









#69
Atrazine
Concen: 40.48 ng
RT: 11.127 min Scan# 1537
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:200 Resp: 446761
Ion Ratio Lower Upper
200 100
173 28.1 6.6 46.6
215 48.8 28.1 68.1

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#70
Pentachlorophenol
Concen: 91.47 ng
RT: 11.233 min Scan# 1555
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:266 Resp: 750925
Ion Ratio Lower Upper
266 100
268 63.4 50.5 75.7
264 63.9 49.8 74.8









#71
Phenanthrene
Concen: 38.68 ng
RT: 11.457 min Scan# 1593
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion:178 Resp: 2663726
Ion Ratio Lower Upper
178 100
176 20.9 16.2 24.2
179 16.5 12.7 19.1

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#72
Anthracene
Concen: 41.65 ng
RT: 11.510 min Scan# 1602
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28


Tgt Ion:178 Resp: 2751581
Ion Ratio Lower Upper
178 100
176 20.4 15.6 23.4
179 16.6 12.8 19.2





#73
 Carbazole
 Concen: 40.32 ng
 RT: 11.663 min Scan# 1628
 Delta R.T. -0.000 min
 Lab File: BF123054.D
 Acq: 28 Jan 2021 11:28

Instrument :
 BNA_F
 ClientSampleId :
 PB134359BS






Tgt Ion:167 Resp: 2471900


Ion	Ratio	Lower	Upper
167	100		
166	22.8	18.0	27.0
139	13.8	10.9	16.3

Manual Integrations
 APPROVED

mohammad
 1/29/2021 10:33:26 AM






#74
 Di-n-butylphthalate
 Concen: 40.73 ng
 RT: 11.992 min Scan# 1684
 Delta R.T. -0.006 min
 Lab File: BF123054.D
 Acq: 28 Jan 2021 11:28



Tgt Ion:149 Resp: 2960967


Ion	Ratio	Lower	Upper
149	100		
150	10.1	7.8	11.6
104	5.5	4.2	6.4





#75
Fluoranthene
Concen: 40.27 ng
RT: 12.651 min Scan# 1796
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Instrument :
BNA_F
ClientSampleId :
PB134359BS




Tgt Ion:202 Resp: 2771716
Ion Ratio Lower Upper
202 100
101 13.6 0.0 32.2
203 18.6 0.0 38.1

Manual Integrations
APPROVED



mohammad
1/29/2021 10:33:26 AM




#76
Chrysene-d12
Concen: 20.00 ng
RT: 14.086 min Scan# 2040
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion:240 Resp: 893039
Ion Ratio Lower Upper
240 100
120 11.9 8.6 12.8
236 25.8 20.3 30.5





#77
Benzidine
Concen: 53.28 ng
RT: 12.768 min Scan# 1816
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Instrument :
BNA_F
ClientSampleId :
PB134359BS



Tgt Ion:184 Resp: 1076747
Ion Ratio Lower Upper
184 100
185 15.0 11.8 17.6
183 11.9 9.7 14.5




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#78
Pyrene
Concen: 41.50 ng
RT: 12.880 min Scan# 1835
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28


Tgt Ion:202 Resp: 2797694
Ion Ratio Lower Upper
202 100
200 23.2 17.6 26.4
203 19.1 14.6 21.8





#79
 Terphenyl-d14
 Concen: 71.77 ng
 RT: 13.021 min Scan# 1859
 Delta R.T. -0.006 min
 Lab File: BF123054.D
 Acq: 28 Jan 2021 11:28




Instrument :
 BNA_F
 ClientSampleId :
 PB134359BS




Tgt Ion:244 Resp: 3262073
 Ion Ratio Lower Upper
 244 100
 212 7.5 5.7 8.5
 122 11.8 8.3 12.5

Manual Integrations
 APPROVED



mohammad
 1/29/2021 10:33:26 AM




#80
 Butylbenzylphthalate
 Concen: 45.06 ng
 RT: 13.498 min Scan# 1940
 Delta R.T. -0.006 min
 Lab File: BF123054.D
 Acq: 28 Jan 2021 11:28




Tgt Ion:149 Resp: 1307108
 Ion Ratio Lower Upper
 149 100
 91 75.1 56.5 84.7
 206 21.1 17.0 25.6





#81
Benzo(a)anthracene
Concen: 41.18 ng
RT: 14.074 min Scan# 2038
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Instrument :
BNA_F
ClientSampled :
PB134359BS




Tgt Ion:228 Resp: 2488677
Ion Ratio Lower Upper
228 100
226 28.8 23.0 34.4
229 20.7 16.6 24.8

Manual Integrations
APPROVED



mohammad
1/29/2021 10:33:26 AM




#82
3,3'-Dichlorobenzidine
Concen: 41.20 ng
RT: 14.033 min Scan# 2031
Delta R.T. 0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28



Tgt Ion:252 Resp: 783685
Ion Ratio Lower Upper
252 100
254 64.5 52.7 79.1
126 12.7 8.0 12.0#









#83
Chrysene
Concen: 40.50 ng
RT: 14.115 min Scan# 2045
Delta R.T. 0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Instrument :
BNA_F
ClientSampleId :
PB134359BS

Tgt Ion:228 Resp: 2449739
Ion Ratio Lower Upper
228 100
226 32.3 24.6 36.8
229 21.0 15.8 23.6




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM



#84
Bis(2-ethylhexyl)phthalate
Concen: 43.46 ng
RT: 14.062 min Scan# 2036
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28


Tgt Ion:149 Resp: 1675232
Ion Ratio Lower Upper
149 100
167 27.8 21.4 32.2
279 4.5 3.4 5.0





#85
Di-n-octyl phthalate
Concen: 46.31 ng
RT: 14.680 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Instrument :
BNA_F
ClientSampleId :
PB134359BS




Tgt Ion:149 Resp: 2998143
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 10.3 7.6 11.4

Manual Integrations
APPROVED



mohammad
1/29/2021 10:33:26 AM




#86
Perylene-d12
Concen: 20.00 ng
RT: 15.568 min Scan# 2292
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28



Tgt Ion:264 Resp: 815560
Ion Ratio Lower Upper
264 100
260 24.3 18.6 28.0
265 21.9 17.0 25.6









#87
Indeno(1,2,3-cd)pyrene
Concen: 44.59 ng
RT: 17.074 min Scan# 2548
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion: 276 Resp: 2421451
Ion Ratio Lower Upper
276 100
138 27.4 18.7 28.1
277 25.1 20.2 30.4

Instrument :
BNA_F
ClientSampleId :
PB134359BS




Manual Integrations
APPROVED


mohammad
1/29/2021 10:33:26 AM




#88
Benzo(b)fluoranthene
Concen: 45.24 ng
RT: 15.139 min Scan# 2219
Delta R.T. 0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28

Tgt Ion: 252 Resp: 2663031
Ion Ratio Lower Upper
252 100
253 22.9 18.3 27.5
125 10.3 8.1 12.1









#89
Benzo(k)fluoranthene
Concen: 41.73 ng
RT: 15.168 min Scan# 2224
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28





Tgt Ion:252 Resp: 2282705
Ion Ratio Lower Upper
252 100
253 23.1 17.8 26.8
125 10.8 8.4 12.6



#90
Benzo(a)pyrene
Concen: 43.00 ng
RT: 15.509 min Scan# 2282
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion:252 Resp: 2223047
Ion Ratio Lower Upper
252 100
253 22.5 17.5 26.3
125 12.3 8.6 12.8




Instrument :
BNA_F
ClientSampleId :
PB134359BS

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM






#91
Dibenzo(a,h)anthracene
Concen: 44.89 ng
RT: 17.098 min Scan# 2552
Delta R.T. -0.000 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28




Tgt Ion: 278 Resp: 2006730

Ion	Ratio	Lower	Upper
278	100		
139	17.6	12.0	18.0
279	24.3	19.4	29.2





#92
Benzo(g,h,i)perylene
Concen: 45.08 ng
RT: 17.533 min Scan# 2626
Delta R.T. 0.006 min
Lab File: BF123054.D
Acq: 28 Jan 2021 11:28



Tgt Ion: 276 Resp: 1952765

Ion	Ratio	Lower	Upper
276	100		
277	24.1	19.3	28.9
138	24.4	16.6	24.8



Instrument :
BNA_F
ClientSampleId :
PB134359BS

Manual Integrations
APPROVED

mohammad
1/29/2021 10:33:26 AM