

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF012918\
 Data File : BF102623.D
 Acq On : 30 Jan 2018 3:21
 Operator : SJ/JU
 Sample : J1353-09
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EPE-106-5(80-82)

Quant Time: Jan 30 05:14:23 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF012218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 20:48:40 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.71	152	174041	20.00	ng	0.00
21) Naphthalene-d8	7.99	136	692128	20.00	ng	0.00
38) Acenaphthene-d10	9.75	164	293140	20.00	ng	0.00
63) Phenanthrene-d10	11.23	188	421283	20.00	ng	0.00
75) Chrysene-d12	13.87	240	284678	20.00	ng	0.00
86) Perylene-d12	15.30	264	293803	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	1275654	111.14	ng	0.01
7) Phenol-d6	6.36	99	1601950	115.76	ng	0.00
23) Nitrobenzene-d5	7.28	82	983484	81.66	ng	0.00
41) 2,4,6-Tribromophenol	10.54	330	246729	84.94	ng	0.00
44) 2-Fluorobiphenyl	9.07	172	1447360	72.60	ng	0.00
78) Terphenyl-d14	12.82	244	889733	70.46	ng	0.00
Target Compounds						
10) Phenol	6.37	94	76711	5.078	ng	96
49) Dimethylphthalate	9.47	163	282615	11.664	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed