

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012920\
 Data File : BF118740.D
 Acq On : 29 Jan 2020 19:17
 Operator : CG/JU
 Sample : L1287-05
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AOC-3-EXC-PIT-(0-5)-A

Quant Time: Jan 30 01:10:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 27 14:24:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	290375	20.00	ng	-0.01
21) Naphthalene-d8	8.13	136	1073174	20.00	ng	-0.02
39) Acenaphthene-d10	9.88	164	587609	20.00	ng	-0.02
64) Phenanthrene-d10	11.36	188	1019116	20.00	ng	-0.02
76) Chrysene-d12	14.00	240	700953	20.00	ng	-0.02
87) Perylene-d12	15.46	264	765640	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	1754466	112.22	ng	0.00
7) Phenol-d6	6.47	99	1900714	93.53	ng	-0.02
23) Nitrobenzene-d5	7.41	82	1568056	81.73	ng	-0.01
42) 2,4,6-Tribromophenol	10.67	330	829520	122.05	ng	-0.02
45) 2-Fluorobiphenyl	9.20	172	2993134	83.72	ng	-0.02
79) Terphenyl-d14	12.95	244	3337541	103.85	ng	-0.01
Target Compounds						
31) Naphthalene	8.15	128	1119418	23.091	ng	98
37) 2-Methylnaphthalene	8.84	142	403837	11.801	ng	98
38) 1-Methylnaphthalene	8.94	142	570770	17.557	ng	99
50) Dimethylphthalate	9.59	163	197784	5.166	ng	99
52) Acenaphthene	9.92	154	239355	7.454	ng	99
58) Fluorene	10.43	166	127805	3.659	ng	98
71) Phenanthrene	11.39	178	185259	3.768	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012920\
Data File : BF118740.D
Acq On : 29 Jan 2020 19:17
Operator : CG/JU
Sample : L1287-05
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
AOC-3-EXC-PIT-(0-5)-A

Quant Time: Jan 30 01:10:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 27 14:24:10 2020
Response via : Initial Calibration

