

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012921\
 Data File : BF123085.D
 Acq On : 29 Jan 2021 16:42
 Operator : JU/CG
 Sample : M1255-12
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CRT-17

Quant Time: Jan 29 23:05:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 29 12:29:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.893	152	252603	20.00	ng	0.00
21) Naphthalene-d8	8.175	136	976379	20.00	ng	0.00
39) Acenaphthene-d10	9.934	164	521368	20.00	ng	0.00
64) Phenanthrene-d10	11.428	188	948932	20.00	ng	0.00
76) Chrysene-d12	14.074	240	805065	20.00	ng	-0.01
86) Perylene-d12	15.563	264	635375	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1838260	112.26	ng	0.02
7) Phenol-d6	6.516	99	2178582	98.15	ng	0.00
23) Nitrobenzene-d5	7.451	82	1519104	78.30	ng	-0.01
42) 2,4,6-Tribromophenol	10.728	330	764034	135.00	ng	0.00
45) 2-Fluorobiphenyl	9.257	172	2622045	74.76	ng	0.00
79) Terphenyl-d14	13.027	244	3143720	76.72	ng	0.00
Target Compounds						
						Qvalue
32) Benzoic acid	7.857	122	4989	5.29	ng	95
50) Dimethylphthalate	9.645	163	110847	2.71	ng	99
52) Acenaphthene	9.963	154	148407	4.41	ng	100
55) Dibenzofuran	10.139	168	120206	2.45	ng	97
58) Fluorene	10.481	166	165893	4.24	ng	97
71) Phenanthrene	11.451	178	438935	7.45	ng	96
75) Fluoranthene	12.639	202	123839	2.10	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012921\
Data File : BF123085.D
Acq On : 29 Jan 2021 16:42
Operator : JU/CG
Sample : M1255-12
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CRT-17

Quant Time: Jan 29 23:05:54 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 29 12:29:00 2021
Response via : Initial Calibration

