

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012921\
 Data File : BF123073.D
 Acq On : 29 Jan 2021 10:09
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 29 12:25:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 28 00:48:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.645	0.649	-0.6	98	0.00
3	Pyridine	1.725	1.731	-0.3	95	0.00
4	n-Nitrosodimethylamine	0.726	0.714	1.7	92	0.00
5 S	2-Fluorophenol	1.297	1.296	0.1	96	0.00
6	Aniline	2.163	2.155	0.4	94	0.00
7 S	Phenol-d6	1.757	1.729	1.6	94	0.00
8	2-Chlorophenol	1.405	1.430	-1.8	97	0.00
9	Benzaldehyde	0.803	0.850	-5.9	100	0.00
10 C	Phenol	1.743	1.861	-6.8	100	0.00
11	bis(2-Chloroethyl)ether	1.450	1.421	2.0	93	0.00
12	1,3-Dichlorobenzene	1.639	1.648	-0.5	96	0.00
13 C	1,4-Dichlorobenzene	1.707	1.649	3.4	96	0.00
14	1,2-Dichlorobenzene	1.597	1.528	4.3	95	0.00
15	Benzyl Alcohol	1.232	1.219	1.1	91	0.00
16	2,2'-oxybis(1-Chloropropane	2.237	2.200	1.7	93	0.00
17	2-Methylphenol	1.196	1.212	-1.3	95	0.00
18	Hexachloroethane	0.598	0.609	-1.8	96	0.00
19 P	n-Nitroso-di-n-propylamine	1.057	1.029	2.6	92	0.00
20	3+4-Methylphenols	1.409	1.466	-4.0	95	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
22	Acetophenone	0.521	0.507	2.7	93	0.00
23 S	Nitrobenzene-d5	0.397	0.413	-4.0	97	0.00
24	Nitrobenzene	0.409	0.425	-3.9	96	0.00
25	Isophorone	0.728	0.726	0.3	94	0.00
26 C	2-Nitrophenol	0.169	0.197	-16.6	100	0.00
27	2,4-Dimethylphenol	0.306	0.299	2.3	92	0.00
28	bis(2-Chloroethoxy)methane	0.497	0.493	0.8	94	0.00
29 C	2,4-Dichlorophenol	0.319	0.334	-4.7	97	0.00
30	1,2,4-Trichlorobenzene	0.360	0.366	-1.7	96	0.00
31	Naphthalene	1.097	1.095	0.2	96	0.00
32	Benzoic acid	0.224	0.267	-19.2	104	0.00
33	4-Chloroaniline	0.487	0.481	1.2	94	0.00
34 C	Hexachlorobutadiene	0.208	0.215	-3.4	97	0.00
35	Caprolactam	0.108	0.116	-7.4	100	0.00
36 C	4-Chloro-3-methylphenol	0.332	0.349	-5.1	97	0.00
37	2-Methylnaphthalene	0.728	0.731	-0.4	96	0.00
38	1-Methylnaphthalene	0.690	0.686	0.6	94	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	0.622	0.637	-2.4	97	0.00
41 P	Hexachlorocyclopentadiene	0.295	0.306	-3.7	92	0.00
42 S	2,4,6-Tribromophenol	0.217	0.243	-12.0	104	0.00
43 C	2,4,6-Trichlorophenol	0.427	0.457	-7.0	98	0.00
44	2,4,5-Trichlorophenol	0.446	0.465	-4.3	96	0.00
45 S	2-Fluorobiphenyl	1.345	1.333	0.9	96	0.00
46	1,1'-Biphenyl	1.640	1.630	0.6	94	0.00
47	2-Chloronaphthalene	1.373	1.364	0.7	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012921\
 Data File : BF123073.D
 Acq On : 29 Jan 2021 10:09
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 29 12:25:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 28 00:48:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.416	0.460	-10.6	99	0.00
49	Acenaphthylene	2.010	2.024	-0.7	96	0.00
50	Dimethylphthalate	1.569	1.596	-1.7	98	0.00
51	2,6-Dinitrotoluene	0.334	0.365	-9.3	98	0.00
52 C	Acenaphthene	1.291	1.308	-1.3	96	0.00
53	3-Nitroaniline	0.395	0.427	-8.1	99	0.00
54 P	2,4-Dinitrophenol	0.126	0.161	-27.8#	112	0.00
55	Dibenzofuran	1.879	1.831	2.6	96	0.00
56 P	4-Nitrophenol	0.286	0.325	-13.6	103	0.00
57	2,4-Dinitrotoluene	0.435	0.486	-11.7	100	0.00
58	Fluorene	1.501	1.379	8.1	97	0.00
59	2,3,4,6-Tetrachlorophenol	0.378	0.413	-9.3	101	0.00
60	Diethylphthalate	1.543	1.575	-2.1	97	0.00
61	4-Chlorophenyl-phenylether	0.705	0.693	1.7	98	0.00
62	4-Nitroaniline	0.397	0.426	-7.3	99	0.00
63	Azobenzene	1.510	1.499	0.7	94	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.101	0.117	-15.8	107	0.00
66 c	n-Nitrosodiphenylamine	0.721	0.703	2.5	96	0.00
67	4-Bromophenyl-phenylether	0.255	0.261	-2.4	99	0.00
68	Hexachlorobenzene	0.273	0.282	-3.3	100	0.00
69	Atrazine	0.199	0.203	-2.0	101	0.00
70 C	Pentachlorophenol	0.148	0.167	-12.8	104	0.00
71	Phenanthrene	1.242	1.178	5.2	98	0.00
72	Anthracene	1.192	1.163	2.4	97	0.00
73	Carbazole	1.106	1.097	0.8	98	0.00
74	Di-n-butylphthalate	1.311	1.311	0.0	98	0.00
75 C	Fluoranthene	1.241	1.216	2.0	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
77	Benzydine	0.453	0.534	-17.9	97	0.00
78	Pyrene	1.510	1.490	1.3	98	0.00
79 S	Terphenyl-d14	1.018	1.044	-2.6	100	0.00
80	Butylbenzylphthalate	0.650	0.697	-7.2	102	0.00
81	Benzo(a)anthracene	1.354	1.340	1.0	99	0.00
82	3,3'-Dichlorobenzidine	0.426	0.476	-11.7	105	0.00
83	Chrysene	1.355	1.347	0.6	101	0.00
84	Bis(2-ethylhexyl)phthalate	0.863	0.890	-3.1	100	0.00
85 c	Di-n-octyl phthalate	1.450	1.578	-8.8	101	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	1.332	1.411	-5.9	103	0.00
88	Benzo(b)fluoranthene	1.444	1.510	-4.6	102	0.00
89	Benzo(k)fluoranthene	1.341	1.333	0.6	102	0.00
90 C	Benzo(a)pyrene	1.268	1.322	-4.3	101	0.00
91	Dibenzo(a,h)anthracene	1.096	1.150	-4.9	101	0.00
92	Benzo(g,h,i)perylene	1.062	1.092	-2.8	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012921\
Data File : BF123073.D
Acq On : 29 Jan 2021 10:09
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 29 12:25:53 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 28 00:48:07 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0