

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012924\
 Data File : BF137184.D
 Acq On : 29 Jan 2024 10:33
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 29 10:56:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 02:14:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	111	0.00
2	1,4-Dioxane	40.000	38.985	2.5	104	0.01
3	Pyridine	40.000	40.714	-1.8	112	0.01
4	n-Nitrosodimethylamine	40.000	37.276	6.8	103	0.01
5 S	2-Fluorophenol	80.000	88.105	-10.1	105	0.00
6	Aniline	40.000	42.559	-6.4	115	0.00
7 S	Phenol-d6	80.000	86.691	-8.4	107	0.00
8	2-Chlorophenol	40.000	42.528	-6.3	109	0.00
9	Benzaldehyde	40.000	46.768	-16.9	105	0.00
10 C	Phenol	40.000	42.954	-7.4	106	0.00
11	bis(2-Chloroethyl)ether	40.000	39.337	1.7	111	0.00
12	1,3-Dichlorobenzene	40.000	40.507	-1.3	108	0.00
13 C	1,4-Dichlorobenzene	40.000	41.556	-3.9	109	0.00
14	1,2-Dichlorobenzene	40.000	40.052	-0.1	106	0.00
15	Benzyl Alcohol	40.000	42.889	-7.2	110	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.116	2.2	105	0.00
17	2-Methylphenol	40.000	42.081	-5.2	105	0.00
18	Hexachloroethane	40.000	40.136	-0.3	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.919	-2.3	108	0.00
20	3+4-Methylphenols	40.000	43.417	-8.5	108	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	41.378	-3.4	110	0.00
23 S	Nitrobenzene-d5	80.000	79.896	0.1	105	0.00
24	Nitrobenzene	40.000	40.414	-1.0	105	0.00
25	Isophorone	40.000	40.910	-2.3	112	0.00
26 C	2-Nitrophenol	40.000	32.387	19.0	87	0.00
27	2,4-Dimethylphenol	40.000	43.560	-8.9	113	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.079	-7.7	113	0.00
29 C	2,4-Dichlorophenol	40.000	45.192	-13.0	111	0.00
30	1,2,4-Trichlorobenzene	40.000	40.966	-2.4	110	0.00
31	Naphthalene	40.000	40.804	-2.0	109	0.00
32	Benzoic acid	40.000	41.967	-4.9	119	0.00
33	4-Chloroaniline	40.000	46.424	-16.1	111	0.00
34 C	Hexachlorobutadiene	40.000	40.513	-1.3	110	0.00
35	Caprolactam	40.000	42.911	-7.3	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.182	-8.0	113	0.00
37	2-Methylnaphthalene	40.000	41.414	-3.5	110	0.00
38	1-Methylnaphthalene	40.000	40.616	-1.5	111	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.014	-0.0	110	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.849	2.9	100	0.00
42 S	2,4,6-Tribromophenol	80.000	80.729	-0.9	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.018	-5.0	109	0.00
44	2,4,5-Trichlorophenol	40.000	44.188	-10.5	117	0.00
45 S	2-Fluorobiphenyl	80.000	78.881	1.4	110	0.00
46	1,1'-Biphenyl	40.000	40.285	-0.7	110	0.00
47	2-Chloronaphthalene	40.000	39.925	0.2	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012924\
 Data File : BF137184.D
 Acq On : 29 Jan 2024 10:33
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 29 10:56:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 02:14:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.909	-2.3	100	0.00
49	Acenaphthylene	40.000	40.333	-0.8	111	0.00
50	Dimethylphthalate	40.000	39.954	0.1	108	0.00
51	2,6-Dinitrotoluene	40.000	39.002	2.5	101	0.00
52 C	Acenaphthene	40.000	39.544	1.1	110	0.00
53	3-Nitroaniline	40.000	38.687	3.3	104	0.00
54 P	2,4-Dinitrophenol	40.000	35.562	11.1	97	0.00
55	Dibenzofuran	40.000	40.027	-0.1	109	0.00
56 P	4-Nitrophenol	40.000	38.427	3.9	110	0.00
57	2,4-Dinitrotoluene	40.000	39.260	1.9	98	0.00
58	Fluorene	40.000	39.354	1.6	110	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.109	-2.8	109	0.00
60	Diethylphthalate	40.000	40.430	-1.1	110	0.00
61	4-Chlorophenyl-phenylether	40.000	39.299	1.8	108	0.00
62	4-Nitroaniline	40.000	40.976	-2.4	104	0.00
63	Azobenzene	40.000	40.061	-0.2	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	40.000	32.268	19.3	87	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.253	-3.1	109	0.00
67	4-Bromophenyl-phenylether	40.000	41.987	-5.0	111	0.00
68	Hexachlorobenzene	40.000	39.999	0.0	107	0.00
69	Atrazine	40.000	42.330	-5.8	110	0.00
70 C	Pentachlorophenol	40.000	41.030	-2.6	107	0.00
71	Phenanthrene	40.000	40.543	-1.4	107	0.00
72	Anthracene	40.000	40.984	-2.5	110	0.00
73	Carbazole	40.000	41.506	-3.8	107	0.00
74	Di-n-butylphthalate	40.000	42.338	-5.8	109	0.00
75 C	Fluoranthene	40.000	40.319	-0.8	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzydine	40.000	44.098	-10.2	127	0.00
78	Pyrene	40.000	40.993	-2.5	107	0.00
79 S	Terphenyl-d14	80.000	81.341	-1.7	107	0.00
80	Butylbenzylphthalate	40.000	39.151	2.1	108	0.00
81	Benzo(a)anthracene	40.000	41.844	-4.6	113	0.00
82	3,3'-Dichlorobenzidine	40.000	44.788	-12.0	113	0.00
83	Chrysene	40.000	40.827	-2.1	110	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.438	-13.6	117	0.00
85 c	Di-n-octyl phthalate	40.000	44.155	-10.4	119	0.00
86 I	Perylene-d12	20.000	20.000	0.0	112	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.872	-2.2	102	0.00
88	Benzo(b)fluoranthene	40.000	39.050	2.4	109	0.00
89	Benzo(k)fluoranthene	40.000	41.623	-4.1	114	0.00
90 C	Benzo(a)pyrene	40.000	40.370	-0.9	109	0.00
91	Dibenzo(a,h)anthracene	40.000	40.942	-2.4	102	0.00
92	Benzo(g,h,i)perylene	40.000	40.338	-0.8	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012924\
Data File : BF137184.D
Acq On : 29 Jan 2024 10:33
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 29 10:56:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 27 02:14:33 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0