

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012925\
 Data File : BF141298.D
 Acq On : 29 Jan 2025 17:31
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF012925

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/30/2025
 Supervised By :mohammad ahmed 02/04/2025

Quant Time: Jan 30 03:16:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 29 17:41:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	133318	20.000	ng	0.00
21) Naphthalene-d8	8.081	136	525644	20.000	ng	0.00
39) Acenaphthene-d10	9.839	164	279764	20.000	ng	0.00
64) Phenanthrene-d10	11.322	188	450291	20.000	ng	0.00
76) Chrysene-d12	13.969	240	257177	20.000	ng	0.00
86) Perylene-d12	15.421	264	279455	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.404	112	650985	74.963	ng	0.00
7) Phenol-d6	6.428	99	824574	74.641	ng	-0.01
23) Nitrobenzene-d5	7.363	82	754118	75.630	ng	0.00
42) 2,4,6-Tribromophenol	10.628	330	192773	75.161	ng	0.00
45) 2-Fluorobiphenyl	9.157	172	1325725	72.938	ng	0.00
79) Terphenyl-d14	12.910	244	1208358	72.460	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.487	88	142440	37.296	ng	100
3) Pyridine	3.216	79	353478	38.413	ng	100
4) n-Nitrosodimethylamine	3.169	42	199637	37.977	ng	100
6) Aniline	6.463	93	368792	39.437	ng	99
8) 2-Chlorophenol	6.581	128	349085	37.505	ng	99
9) Benzaldehyde	6.346	77	236537	36.965	ng	100
10) Phenol	6.440	94	443695	37.888	ng	99
11) bis(2-Chloroethyl)ether	6.534	93	335702	37.387	ng	100
12) 1,3-Dichlorobenzene	6.740	146	382682	37.288	ng	99
13) 1,4-Dichlorobenzene	6.816	146	390597	37.885	ng	99
14) 1,2-Dichlorobenzene	6.969	146	359529	37.241	ng	100
15) Benzyl Alcohol	6.940	79	286519	37.952	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.075	45	587637	37.636	ng	100
17) 2-Methylphenol	7.051	107	276450	36.544	ng	99
18) Hexachloroethane	7.310	117	144890	38.094	ng	99
19) n-Nitroso-di-n-propyla...	7.210	70	241878	36.848	ng	100
20) 3+4-Methylphenols	7.204	107	347046	37.364	ng	99
22) Acetophenone	7.210	105	458757	36.717	ng	100
24) Nitrobenzene	7.381	77	391907	37.927	ng	99
25) Isophorone	7.622	82	645309	37.439	ng	100
26) 2-Nitrophenol	7.698	139	186592	38.289	ng	99
27) 2,4-Dimethylphenol	7.734	122	236052m	38.055	ng	
28) bis(2-Chloroethoxy)met...	7.834	93	411851	37.460	ng	99
29) 2,4-Dichlorophenol	7.940	162	286090	37.675	ng	99
30) 1,2,4-Trichlorobenzene	8.022	180	316431	36.488	ng	100
31) Naphthalene	8.104	128	1030995	37.128	ng	100
32) Benzoic acid	7.845	122	236833	41.534	ng	99
33) 4-Chloroaniline	8.151	127	350423	37.214	ng	99
34) Hexachlorobutadiene	8.222	225	186523	36.676	ng	99
35) Caprolactam	8.522	113	93403	37.662	ng	97
36) 4-Chloro-3-methylphenol	8.634	107	305288	37.480	ng	99
37) 2-Methylnaphthalene	8.792	142	651222	36.898	ng	99
38) 1-Methylnaphthalene	8.892	142	651341	37.570	ng	99
40) 1,2,4,5-Tetrachloroben...	8.957	216	296277	37.226	ng	99
41) Hexachlorocyclopentadiene	8.945	237	93930	39.961	ng	98
43) 2,4,6-Trichlorophenol	9.069	196	201214	38.625	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012925\
 Data File : BF141298.D
 Acq On : 29 Jan 2025 17:31
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF012925

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/30/2025
 Supervised By :mohammad ahmed 02/04/2025

Quant Time: Jan 30 03:16:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 29 17:41:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.110	196	209611	36.969	ng	100
46) 1,1'-Biphenyl	9.257	154	809694	37.111	ng	100
47) 2-Chloronaphthalene	9.281	162	626457	36.837	ng	100
48) 2-Nitroaniline	9.381	65	203104	37.970	ng	98
49) Acenaphthylene	9.698	152	890709	37.416	ng	100
50) Dimethylphthalate	9.563	163	697300	36.904	ng	100
51) 2,6-Dinitrotoluene	9.622	165	164343	37.459	ng	99
52) Acenaphthene	9.869	154	632756	37.206	ng	99
53) 3-Nitroaniline	9.792	138	164042	38.150	ng	98
54) 2,4-Dinitrophenol	9.892	184	81034	39.734	ng	# 85
55) Dibenzofuran	10.045	168	842555	37.012	ng	100
56) 4-Nitrophenol	9.945	139	122925	39.089	ng	99
57) 2,4-Dinitrotoluene	10.022	165	217705	38.327	ng	99
58) Fluorene	10.386	166	643639	36.451	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.163	232	169067	37.902	ng	99
60) Diethylphthalate	10.263	149	683674	37.316	ng	99
61) 4-Chlorophenyl-phenyle...	10.381	204	312065	36.180	ng	99
62) 4-Nitroaniline	10.404	138	159227	37.665	ng	99
63) Azobenzene	10.539	77	678905	37.020	ng	100
65) 4,6-Dinitro-2-methylph...	10.433	198	111635	41.913	ng	99
66) n-Nitrosodiphenylamine	10.498	169	564452	38.757	ng	99
67) 4-Bromophenyl-phenylether	10.869	248	199009	39.477	ng	98
68) Hexachlorobenzene	10.933	284	208171	38.929	ng	99
69) Atrazine	11.028	200	190958	39.215	ng	99
70) Pentachlorophenol	11.128	266	128300	40.973	ng	99
71) Phenanthrene	11.351	178	922625	38.117	ng	100
72) Anthracene	11.398	178	912703	38.484	ng	100
73) Carbazole	11.557	167	813036	38.661	ng	100
74) Di-n-butylphthalate	11.892	149	984759	38.639	ng	100
75) Fluoranthene	12.539	202	902637	37.768	ng	100
77) Benzidine	12.663	184	405163	49.142	ng	99
78) Pyrene	12.769	202	894269	36.225	ng	99
80) Butylbenzylphthalate	13.392	149	324793	37.660	ng	99
81) Benzo(a)anthracene	13.957	228	631556	35.587	ng	99
82) 3,3'-Dichlorobenzidine	13.921	252	218850	38.774	ng	99
83) Chrysene	13.992	228	610038	37.508	ng	100
84) Bis(2-ethylhexyl)phtha...	13.951	149	400642	36.620	ng	100
85) Di-n-octyl phthalate	14.569	149	642439	37.807	ng	99
87) Indeno(1,2,3-cd)pyrene	16.868	276	678653	37.626	ng	99
88) Benzo(b)fluoranthene	14.998	252	683111	38.861	ng	100
89) Benzo(k)fluoranthene	15.027	252	556960	35.766	ng	99
90) Benzo(a)pyrene	15.363	252	567574	38.209	ng	99
91) Dibenzo(a,h)anthracene	16.886	278	551534	38.007	ng	99
92) Benzo(g,h,i)perylene	17.304	276	559807	37.101	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

