


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF013016\
 Data File : BF084693.D
 Acq On : 30 Jan 2016 19:26
 Operator : SJ/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF012916.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 29 12:54:53 2016



AutoFind: Scans 824, 825, 826; Background Corrected with Scan 821

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.7	16711	PASS
68	69	0.00	2	1.6	238	PASS
69	198	0.00	100	35.4	15288	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.9	21120	PASS
197	198	0.00	2	0.4	185	PASS
198	198	100	100	100.0	43191	PASS
199	198	5	9	6.8	2953	PASS
275	198	10	60	24.5	10594	PASS
365	198	1	100	2.6	1140	PASS
441	442	0.01	24	16.3	6250	PASS
442	198	50	100	88.8	38357	PASS
443	442	15	24	19.4	7448	PASS