

Data Path : Z:\HPCHEM1\BNA F\DATA\BF013017\
 Data File : BF092641.D
 Acq On : 30 Jan 2017 18:39
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 31 01:14:53 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 30 14:50:05 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.630	0.743	-17.9	127	0.01
3	Pyridine	1.602	2.067	-29.0#	134	0.00
4	n-Nitrosodimethylamine	0.752	0.952	-26.6#	133	0.01
5 S	2-Fluorophenol	1.166	1.223	-4.9	106	-0.01
6	Aniline	2.046	2.079	-1.6	110	0.00
7 S	Phenol-d6	1.500	1.516	-1.1	107	0.00
8	2-Chlorophenol	1.286	1.321	-2.7	108	0.00
9	Benzaldehyde	1.075	1.133	-5.4	109	0.00
10 C	Phenol	1.755	1.678	4.4	106	0.00
11	bis(2-Chloroethyl)ether	1.401	1.388	0.9	109	0.00
12	1,3-Dichlorobenzene	1.506	1.496	0.7	107	0.00
13 C	1,4-Dichlorobenzene	1.525	1.528	-0.2	109	0.00
14	1,2-Dichlorobenzene	1.458	1.536	-5.3	110	0.00
15	Benzyl Alcohol	1.110	1.103	0.6	108	0.00
16	2,2'-oxybis(1-Chloropropane	2.434	2.410	1.0	106	0.00
17	2-Methylphenol	1.103	1.203	-9.1	110	0.00
18	Hexachloroethane	0.520	0.540	-3.8	109	0.00
19 P	n-Nitroso-di-n-propylamine	0.955	0.906	5.1	110	0.00
20	3+4-Methylphenols	1.319	1.373	-4.1	107	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
22	Acetophenone	0.457	0.429	6.1	107	0.00
23 S	Nitrobenzene-d5	0.359	0.374	-4.2	114	0.00
24	Nitrobenzene	0.369	0.329	10.8	105	0.00
25	Isophorone	0.669	0.647	3.3	109	0.00
26 C	2-Nitrophenol	0.175	0.190	-8.6	110	0.00
27	2,4-Dimethylphenol	0.308	0.321	-4.2	109	0.00
28	bis(2-Chloroethoxy)methane	0.443	0.448	-1.1	112	0.00
29 C	2,4-Dichlorophenol	0.287	0.278	3.1	112	0.01
30	1,2,4-Trichlorobenzene	0.309	0.297	3.9	108	0.00
31	Naphthalene	1.014	1.007	0.7	113	0.00
32	Benzoic acid	0.156	0.164	-5.1	105	0.00
33	4-Chloroaniline	0.398	0.397	0.3	108	0.00
34 C	Hexachlorobutadiene	0.170	0.165	2.9	107	0.00
35	Caprolactam	0.090	0.093	-3.3	106	0.00
36 C	4-Chloro-3-methylphenol	0.299	0.306	-2.3	111	0.00
37	2-Methylnaphthalene	0.651	0.597	8.3	101	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
39	1,2,4,5-Tetrachlorobenzene	0.561	0.582	-3.7	114	0.01
40 P	Hexachlorocyclopentadiene	0.253	0.263	-4.0	112	0.00
41 S	2,4,6-Tribromophenol	0.145	0.156	-7.6	108	0.00
42 C	2,4,6-Trichlorophenol	0.369	0.383	-3.8	107	0.00
43	2,4,5-Trichlorophenol	0.404	0.400	1.0	112	0.00
44 S	2-Fluorobiphenyl	1.104	1.034	6.3	96	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF013017\
 Data File : BF092641.D
 Acq On : 30 Jan 2017 18:39
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 31 01:14:53 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 30 14:50:05 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.448	1.517	-4.8	110	0.00
46	2-Chloronaphthalene	1.133	1.120	1.1	101	0.00
47	2-Nitroaniline	0.381	0.397	-4.2	121	0.01
48	Acenaphthylene	1.957	1.923	1.7	115	0.01
49	Dimethylphthalate	1.347	1.254	6.9	101	0.00
50	2,6-Dinitrotoluene	0.291	0.337	-15.8	123	0.01
51 C	Acenaphthene	1.170	1.166	0.3	115	0.00
52	3-Nitroaniline	0.333	0.391	-17.4	120	0.00
53 P	2,4-Dinitrophenol	0.130	0.150	-15.4	119	0.01
54	Dibenzofuran	1.565	1.533	2.0	114	0.00
55 P	4-Nitrophenol	0.240	0.237	1.3	109	0.00
56	2,4-Dinitrotoluene	0.387	0.425	-9.8	106	0.00
57	Fluorene	1.204	1.169	2.9	109	0.00
58	2,3,4,6-Tetrachlorophenol	0.270	0.296	-9.6	107	0.00
59	Diethylphthalate	1.270	1.229	3.2	102	0.00
60	4-Chlorophenyl-phenylether	0.578	0.582	-0.7	113	0.01
61	4-Nitroaniline	0.341	0.360	-5.6	116	0.00
62	Azobenzene	1.312	1.329	-1.3	111	0.01
63 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
64	4,6-Dinitro-2-methylphenol	0.103	0.133	-29.1#	118	0.00
65 c	n-Nitrosodiphenylamine	0.639	0.617	3.4	102	0.00
66	4-Bromophenyl-phenylether	0.209	0.211	-1.0	111	0.01
67	Hexachlorobenzene	0.206	0.199	3.4	106	0.00
68	Atrazine	0.205	0.191	6.8	106	0.00
69 C	Pentachlorophenol	0.089	0.101	-13.5	114	0.00
70	Phenanthrene	1.146	1.154	-0.7	107	0.00
71	Anthracene	1.151	1.085	5.7	104	0.00
72	Carbazole	1.100	1.147	-4.3	105	0.00
73	Di-n-butylphthalate	1.190	1.161	2.4	105	0.00
74 C	Fluoranthene	1.182	1.160	1.9	103	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	105	-0.01
76	Benzidine	0.852	0.825	3.2	102	0.00
77	Pyrene	1.497	1.542	-3.0	102	0.00
78 S	Terphenyl-d14	0.851	0.838	1.5	108	0.00
79	Butylbenzylphthalate	0.608	0.615	-1.2	106	0.00
80	Benzo(a)anthracene	1.223	1.219	0.3	103	0.00
81	3,3'-Dichlorobenzidine	0.436	0.437	-0.2	102	0.00
82	Chrysene	1.145	1.261	-10.1	107	0.00
83	Bis(2-ethylhexyl)phthalate	0.831	0.899	-8.2	109	0.00
84 c	Di-n-octyl phthalate	1.354	1.508	-11.4	111	0.00
85	Indeno(1,2,3-cd)pyrene	0.887	0.847	4.5	105	0.00
86 I	Perylene-d12	1.000	1.000	0.0	105	0.00
87	Benzo(b)fluoranthene	1.316	1.467	-11.5	111	-0.01

Data Path : Z:\HPCHEM1\BNA F\DATA\BF013017\
Data File : BF092641.D
Acq On : 30 Jan 2017 18:39
Operator : SJ/MA
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 31 01:14:53 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 30 14:50:05 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.137	1.001	12.0	101	0.00
89 C	Benzo(a)pyrene	1.139	1.143	-0.4	106	0.00
90	Dibenzo(a,h)anthracene	0.920	0.866	5.9	105	0.00
91	Benzo(a,h,i)perylene	0.930	0.863	7.2	104	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0