


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF013117\
 Data File : BF092678.D
 Acq On : 31 Jan 2017 13:44
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF013017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 30 14:50:05 2017



AutoFind: Scans 844, 845, 846; Background Corrected with Scan 840

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.5	50506	PASS
68	69	0.00	2	1.9	775	PASS
69	198	0.00	100	36.9	40951	PASS
70	69	0.00	2	0.6	235	PASS
127	198	10	80	46.8	51949	PASS
197	198	0.00	2	0.2	179	PASS
198	198	100	100	100.0	111094	PASS
199	198	5	9	6.7	7460	PASS
275	198	10	60	22.4	24936	PASS
365	198	1	100	2.1	2289	PASS
441	442	0.01	24	16.0	13055	PASS
442	198	50	100	73.5	81698	PASS
443	442	15	24	18.6	15211	PASS