

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013122\
 Data File : BF126746.D
 Acq On : 31 Jan 2022 12:59
 Operator : CG/JU
 Sample : PB142313TB
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB142313TB

Quant Time: Feb 01 02:39:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 27 03:09:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.957	152	161558	20.000	ng	0.00
21) Naphthalene-d8	8.239	136	652988	20.000	ng	0.00
39) Acenaphthene-d10	9.998	164	358080	20.000	ng	0.00
64) Phenanthrene-d10	11.492	188	610737	20.000	ng	0.00
76) Chrysene-d12	14.139	240	447458	20.000	ng	# 0.00
86) Perylene-d12	15.668	264	308632	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.587	112	1442714	117.510	ng	0.02
7) Phenol-d6	6.581	99	1950898	114.369	ng	0.00
23) Nitrobenzene-d5	7.522	82	1399612	83.895	ng	0.00
42) 2,4,6-Tribromophenol	10.792	330	422253	126.005	ng	0.00
45) 2-Fluorobiphenyl	9.322	172	1733189	74.128	ng	0.00
79) Terphenyl-d14	13.086	244	2056401	77.083	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013122\
Data File : BF126746.D
Acq On : 31 Jan 2022 12:59
Operator : CG/JU
Sample : PB142313TB
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB142313TB

Quant Time: Feb 01 02:39:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 27 03:09:40 2022
Response via : Initial Calibration

