

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013123\
 Data File : BF132249.D
 Acq On : 01 Feb 2023 01:32
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 01 06:01:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 05:24:37 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.585	0.588	-0.5	109	0.00
3	Pyridine	1.588	1.556	2.0	105	0.00
4	n-Nitrosodimethylamine	0.508	0.494	2.8	106	0.00
5 S	2-Fluorophenol	1.244	1.206	3.1	107	0.00
6	Aniline	2.082	2.024	2.8	108	0.00
7 S	Phenol-d6	1.534	1.491	2.8	107	0.00
8	2-Chlorophenol	1.289	1.284	0.4	108	0.00
9	Benzaldehyde	0.936	0.844	9.8	107	0.00
10 C	Phenol	1.588	1.535	3.3	106	0.00
11	bis(2-Chloroethyl)ether	1.307	1.289	1.4	108	0.00
12	1,3-Dichlorobenzene	1.363	1.345	1.3	108	0.00
13 C	1,4-Dichlorobenzene	1.361	1.354	0.5	109	0.00
14	1,2-Dichlorobenzene	1.269	1.258	0.9	109	0.00
15	Benzyl Alcohol	1.119	1.108	1.0	107	0.00
16	2,2'-oxybis(1-Chloropropane	1.377	1.332	3.3	105	0.00
17	2-Methylphenol	1.128	1.112	1.4	108	0.00
18	Hexachloroethane	0.510	0.510	0.0	108	0.00
19 P	n-Nitroso-di-n-propylamine	0.871	0.811	6.9	104	0.00
20	3+4-Methylphenols	1.440	1.433	0.5	107	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
22	Acetophenone	0.470	0.453	3.6	106	0.00
23 S	Nitrobenzene-d5	0.359	0.352	1.9	107	0.00
24	Nitrobenzene	0.367	0.358	2.5	106	0.00
25	Isophorone	0.682	0.669	1.9	109	0.00
26 C	2-Nitrophenol	0.179	0.190	-6.1	111	0.00
27	2,4-Dimethylphenol	0.313	0.315	-0.6	110	0.00
28	bis(2-Chloroethoxy)methane	0.419	0.413	1.4	108	0.00
29 C	2,4-Dichlorophenol	0.278	0.280	-0.7	110	0.00
30	1,2,4-Trichlorobenzene	0.298	0.298	0.0	109	0.00
31	Naphthalene	0.984	0.972	1.2	108	0.00
32	Benzoic acid	0.232	0.240	-3.4	110	0.00
33	4-Chloroaniline	0.437	0.435	0.5	108	0.00
34 C	Hexachlorobutadiene	0.167	0.171	-2.4	112	0.00
35	Caprolactam	0.095	0.098	-3.2	112	0.00
36 C	4-Chloro-3-methylphenol	0.299	0.301	-0.7	109	0.00
37	2-Methylnaphthalene	0.667	0.661	0.9	108	0.00
38	1-Methylnaphthalene	0.620	0.618	0.3	109	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
40	1,2,4,5-Tetrachlorobenzene	0.563	0.561	0.4	110	0.00
41 P	Hexachlorocyclopentadiene	0.328	0.348	-6.1	113	0.00
42 S	2,4,6-Tribromophenol	0.206	0.206	0.0	113	0.00
43 C	2,4,6-Trichlorophenol	0.402	0.408	-1.5	110	0.00
44	2,4,5-Trichlorophenol	0.463	0.469	-1.3	111	0.00
45 S	2-Fluorobiphenyl	1.295	1.167	9.9	108	0.00
46	1,1'-Biphenyl	1.542	1.518	1.6	109	0.00
47	2-Chloronaphthalene	1.174	1.149	2.1	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013123\
 Data File : BF132249.D
 Acq On : 01 Feb 2023 01:32
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 01 06:01:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 05:24:37 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.349	0.350	-0.3	107	0.00
49	Acenaphthylene	1.917	1.898	1.0	109	0.00
50	Dimethylphthalate	1.399	1.386	0.9	109	0.00
51	2,6-Dinitrotoluene	0.312	0.325	-4.2	112	0.00
52 C	Acenaphthene	1.200	1.181	1.6	109	0.00
53	3-Nitroaniline	0.379	0.386	-1.8	110	0.00
54 P	2,4-Dinitrophenol	0.161	0.170	-5.6	111	0.00
55	Dibenzofuran	1.686	1.659	1.6	109	0.00
56 P	4-Nitrophenol	0.285	0.297	-4.2	110	0.00
57	2,4-Dinitrotoluene	0.406	0.430	-5.9	113	0.00
58	Fluorene	1.298	1.260	2.9	109	0.00
59	2,3,4,6-Tetrachlorophenol	0.338	0.345	-2.1	109	0.00
60	Diethylphthalate	1.392	1.406	-1.0	110	0.00
61	4-Chlorophenyl-phenylether	0.616	0.603	2.1	109	0.00
62	4-Nitroaniline	0.363	0.364	-0.3	109	0.00
63	Azobenzene	1.356	1.308	3.5	105	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
65	4,6-Dinitro-2-methylphenol	0.114	0.127	-11.4	115	0.00
66 c	n-Nitrosodiphenylamine	0.626	0.611	2.4	108	0.00
67	4-Bromophenyl-phenylether	0.205	0.204	0.5	111	0.00
68	Hexachlorobenzene	0.219	0.218	0.5	111	0.00
69	Atrazine	0.180	0.187	-3.9	110	0.00
70 C	Pentachlorophenol	0.150	0.155	-3.3	110	0.00
71	Phenanthrene	1.045	1.007	3.6	108	0.00
72	Anthracene	1.064	1.035	2.7	108	0.00
73	Carbazole	0.954	0.927	2.8	109	0.00
74	Di-n-butylphthalate	1.127	1.120	0.6	109	0.00
75 C	Fluoranthene	1.067	1.039	2.6	109	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
77	Benzidine	0.699	0.630	9.9	102	0.00
78	Pyrene	1.609	1.592	1.1	108	0.00
79 S	Terphenyl-d14	1.034	0.983	4.9	107	0.00
80	Butylbenzylphthalate	0.711	0.715	-0.6	107	0.00
81	Benzo(a)anthracene	1.399	1.384	1.1	107	0.00
82	3,3'-Dichlorobenzidine	0.446	0.449	-0.7	107	0.00
83	Chrysene	1.310	1.295	1.1	108	0.00
84	Bis(2-ethylhexyl)phthalate	0.961	0.955	0.6	107	0.00
85 c	Di-n-octyl phthalate	1.567	1.587	-1.3	108	0.00
86 I	Perylene-d12	1.000	1.000	0.0	109	0.00
87	Indeno(1,2,3-cd)pyrene	1.329	1.312	1.3	112	0.00
88	Benzo(b)fluoranthene	1.271	1.258	1.0	104	0.00
89	Benzo(k)fluoranthene	1.213	1.264	-4.2	114	0.00
90 C	Benzo(a)pyrene	1.184	1.190	-0.5	109	0.00
91	Dibenzo(a,h)anthracene	1.068	1.065	0.3	111	0.00
92	Benzo(g,h,i)perylene	1.104	1.077	2.4	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013123\
Data File : BF132249.D
Acq On : 01 Feb 2023 01:32
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 01 06:01:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 01 05:24:37 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0