




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013124\
Data File : BF137201.D
Acq On : 31 Jan 2024 10:30
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 01 04:51:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration



TIC: BF137201.D\data.ms

(70) Pentachlorophenol (C)




11.116min (-0.005) 27394.59 ng

response 72328

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.30	65.43
264.00	61.00	62.19
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013124\
Data File : BF137201.D
Acq On : 31 Jan 2024 10:30
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 01 04:51:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration



TIC: BF137201.D\data.ms

(77) Benzidine

12.669min (+ 0.001) 0.00 ng

response 375742

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.21
183.00	11.00	12.15
0.00	0.00	0.00