


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013124\
Data File : BF137201.D
Acq On : 31 Jan 2024 10:30
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Jan 31 04:56:22 2024






AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1609

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.0	32715	PASS
68	69	0.00	2	1.6	543	PASS
69	198	0.00	100	38.4	33056	PASS
70	69	0.00	2	0.2	61	PASS
127	198	10	80	43.0	37035	PASS
197	198	0.00	2	0.4	338	PASS
198	198	100	100	100.0	86115	PASS
199	198	5	9	6.5	5556	PASS
275	198	10	60	26.9	23171	PASS
365	198	1	100	3.2	2760	PASS
441	198	0.01	100	11.9	10281	PASS
442	442	50	100	100.0	71590	PASS
443	442	15	24	18.6	13301	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013124\
Data File : BF137201.D
Acq On : 31 Jan 2024 10:30
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Feb 01 04:51:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration



TIC: BF137201.D\data.ms

(70) Pentachlorophenol (C)

11.116min (-0.005) 27394.59 ng


response 72328

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.30	65.43
264.00	61.00	62.19
0.00	0.00	0.00

```
Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013124\
Data File : BF137201.D
Acq On    : 31 Jan 2024  10:30
Operator  : CG\JU
Sample    : DFTPP
Misc      :
ALS Vial  : 1    Sample Multiplier: 1
```

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Feb 01 04:51:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 04:56:22 2024
Response via : Initial Calibration



(77) Benzidine

12.669min (+ 0.001) 0.00 ng

response **375742**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.21
183.00	11.00	12.15
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/31/2024	BNA_F	<u>BF137201.D</u>
Compound Name	Response	Retention Time
DDT	1505931	13.492
DDD	1871	13.192
DDE	0	12.88
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
1871	1507802	0.12

Instrument :
BNA_F
ClientSampleId :
DFTPP