

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020116\
 Data File : BF084716.D
 Acq On : 1 Feb 2016 18:28
 Operator : SJ/IZ
 Sample : H1283-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S8-0537

Quant Time: Feb 02 01:00:16 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 01 14:55:11 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	161062	20.00	ng	-0.02
21) Naphthalene-d8	8.00	136	685678	20.00	ng	-0.02
38) Acenaphthene-d10	9.74	164	311390	20.00	ng	-0.02
63) Phenanthrene-d10	11.23	188	601020	20.00	ng	-0.01
75) Chrysene-d12	13.86	240	447811	20.00	ng	-0.01
86) Perylene-d12	15.24	264	321216	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	711787	68.84	ng	-0.01
7) Phenol-d6	6.35	99	533197	41.59	ng	-0.02
23) Nitrobenzene-d5	7.28	82	1181126	97.04	ng	-0.02
41) 2,4,6-Tribromophenol	10.54	330	493718	152.00	ng	-0.01
44) 2-Fluorobiphenyl	9.08	172	2028387	138.99	ng	-0.01
78) Terphenyl-d14	12.82	244	1651090	83.55	ng	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020116\
Data File : BF084716.D
Acq On : 1 Feb 2016 18:28
Operator : SJ/IZ
Sample : H1283-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S8-0537

Quant Time: Feb 02 01:00:16 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 01 14:55:11 2016
Response via : Initial Calibration

