

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020117\
 Data File : BF092727.D
 Acq On : 2 Feb 2017 3:04
 Operator : SJ/MA
 Sample : I1615-01
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SS-1

Quant Time: Feb 02 12:23:54 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 30 14:50:05 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.94	152	148757	20.00	ng	-0.02
21) Naphthalene-d8	8.24	136	494831	20.00	ng	-0.02
38) Acenaphthene-d10	10.01	164	232440	20.00	ng	0.00
63) Phenanthrene-d10	11.49	188	342122	20.00	ng	0.00
75) Chrysene-d12	14.12	240	300572	20.00	ng	-0.02
86) Perylene-d12	15.60	264	295311	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	1012697	116.79	ng	-0.01
7) Phenol-d6	6.59	99	1301537	116.66	ng	-0.01
23) Nitrobenzene-d5	7.52	82	796022	89.58	ng	-0.02
41) 2,4,6-Tribromophenol	10.80	330	164671	97.95	ng	-0.01
44) 2-Fluorobiphenyl	9.32	172	1057347	82.41	ng	-0.01
78) Terphenyl-d14	13.07	244	600534	46.95	ng	-0.01
Target Compounds						Qvalue
10) Phenol	6.60	94	86164	6.60	ng	# 65
37) 2-Methylnaphthalene	8.96	142	278418	17.29	ng	99
45) 1,1'-Biphenyl	9.42	154	63296	3.76	ng	96
49) Dimethylphthalate	9.71	163	196242	12.53	ng	# 95
70) Phenanthrene	11.52	178	70165	3.58	ng	# 93
77) Pyrene	12.93	202	48576	2.16	ng	98
83) Bis(2-ethylhexyl)phthalate	14.10	149	29012	2.32	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020117\
Data File : BF092727.D
Acq On : 2 Feb 2017 3:04
Operator : SJ/MA
Sample : I1615-01
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SS-1

Quant Time: Feb 02 12:23:54 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF013017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 30 14:50:05 2017
Response via : Initial Calibration

