


Data Path : Z:\HPCHEM1\BNA F\Data\BF020215\
 Data File : BF077156.D
 Acq On : 2 Feb 2015 12:39 pm
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 30 22:48:24 2015



AutoFind: Scans 1512, 1513, 1514; Background Corrected with Scan 1509

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.6	5080	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	43.2	5406	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.8	6980	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	12502	PASS
199	198	5	9	7.8	973	PASS
275	198	10	60	22.7	2835	PASS
365	198	1	100	1.7	207	PASS
441	442	0.01	24	14.6	1773	PASS
442	198	50	100	97.4	12173	PASS
443	442	15	24	19.4	2357	PASS