

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020216\
 Data File : BF084760.D
 Acq On : 2 Feb 2016 21:55
 Operator : SJ/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 03 02:45:28 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 01 14:55:11 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	-0.03
2	1,4-Dioxane	0.562	0.537	4.4	93	-0.05
3	Pyridine	1.465	1.471	-0.4	101	-0.06
4	n-Nitrosodimethylamine	0.758	0.715	5.7	88	-0.05
5 S	2-Fluorophenol	1.284	1.325	-3.2	104	-0.02
6	Aniline	1.948	1.961	-0.7	97	-0.02
7 S	Phenol-d6	1.592	1.514	4.9	97	-0.02
8	2-Chlorophenol	1.453	1.531	-5.4	99	-0.02
9	Benzaldehyde	0.940	0.922	1.9	91	-0.02
10 C	Phenol	1.783	1.675	6.1	102	-0.02
11	bis(2-Chloroethyl)ether	1.391	1.379	0.9	102	-0.02
12	1,3-Dichlorobenzene	1.536	1.474	4.0	96	-0.02
13 C	1,4-Dichlorobenzene	1.535	1.560	-1.6	102	-0.02
14	1,2-Dichlorobenzene	1.430	1.370	4.2	89	-0.02
15	Benzyl Alcohol	1.099	1.095	0.4	98	-0.02
16	2,2'-oxybis(1-Chloropropane	2.339	2.308	1.3	99	-0.02
17	2-Methylphenol	1.149	1.175	-2.3	92	-0.02
18	Hexachloroethane	0.591	0.578	2.2	92	-0.03
19 P	n-Nitroso-di-n-propylamine	0.998	0.981	1.7	100	-0.02
20	3+4-Methylphenols	1.427	1.381	3.2	94	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	96	-0.03
22	Acetophenone	0.527	0.500	5.1	93	-0.02
23 S	Nitrobenzene-d5	0.355	0.370	-4.2	99	-0.02
24	Nitrobenzene	0.380	0.337	11.3	94	-0.02
25	Isophorone	0.690	0.685	0.7	94	-0.02
26 C	2-Nitrophenol	0.188	0.197	-4.8	102	-0.02
27	2,4-Dimethylphenol	0.326	0.299	8.3	94	-0.02
28	bis(2-Chloroethoxy)methane	0.410	0.388	5.4	94	-0.03
29 C	2,4-Dichlorophenol	0.279	0.278	0.4	92	-0.02
30	1,2,4-Trichlorobenzene	0.315	0.300	4.8	94	-0.03
31	Naphthalene	1.003	0.906	9.7	93	-0.02
32	Benzoic acid	0.209	0.230	-10.0	105	-0.01
33	4-Chloroaniline	0.415	0.388	6.5	99	-0.02
34 C	Hexachlorobutadiene	0.182	0.188	-3.3	98	-0.02
35	Caprolactam	0.086	0.098	-14.0	96	-0.01
36 C	4-Chloro-3-methylphenol	0.318	0.300	5.7	97	-0.02
37	2-Methylnaphthalene	0.628	0.659	-4.9	96	-0.02
38 I	Acenaphthene-d10	1.000	1.000	0.0	93	-0.02
39	1,2,4,5-Tetrachlorobenzene	0.635	0.571	10.1	92	-0.03
40 P	Hexachlorocyclopentadiene	0.223	0.262	-17.5	108	-0.02
41 S	2,4,6-Tribromophenol	0.209	0.218	-4.3	92	-0.02
42 C	2,4,6-Trichlorophenol	0.409	0.402	1.7	90	-0.02
43	2,4,5-Trichlorophenol	0.416	0.405	2.6	94	-0.02
44 S	2-Fluorobiphenyl	1.321	1.313	0.6	104	-0.02

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020216\
 Data File : BF084760.D
 Acq On : 2 Feb 2016 21:55
 Operator : SJ/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 03 02:45:28 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 01 14:55:11 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.786	1.719	3.8	89	-0.02
46	2-Chloronaphthalene	1.316	1.194	9.3	87	-0.03
47	2-Nitroaniline	0.417	0.375	10.1	95	-0.02
48	Acenaphthylene	1.996	1.914	4.1	89	-0.03
49	Dimethylphthalate	1.523	1.572	-3.2	95	-0.02
50	2,6-Dinitrotoluene	0.333	0.369	-10.8	98	-0.02
51 C	Acenaphthene	1.299	1.236	4.8	89	-0.02
52	3-Nitroaniline	0.374	0.339	9.4	87	-0.02
53 P	2,4-Dinitrophenol	0.133	0.159	-19.5	104	-0.01
54	Dibenzofuran	1.587	1.531	3.5	89	-0.02
55 P	4-Nitrophenol	0.275	0.266	3.3	82	-0.01
56	2,4-Dinitrotoluene	0.424	0.466	-9.9	101	-0.02
57	Fluorene	1.326	1.240	6.5	85	-0.02
58	2,3,4,6-Tetrachlorophenol	0.317	0.311	1.9	102	-0.02
59	Diethylphthalate	1.487	1.441	3.1	93	-0.02
60	4-Chlorophenyl-phenylether	0.621	0.611	1.6	98	-0.02
61	4-Nitroaniline	0.364	0.370	-1.6	97	-0.02
62	Azobenzene	1.430	1.517	-6.1	100	-0.02
63 I	Phenanthrene-d10	1.000	1.000	0.0	98	-0.02
64	4,6-Dinitro-2-methylphenol	0.123	0.128	-4.1	96	-0.01
65 c	n-Nitrosodiphenylamine	0.635	0.654	-3.0	94	-0.02
66	4-Bromophenyl-phenylether	0.221	0.233	-5.4	97	-0.02
67	Hexachlorobenzene	0.241	0.224	7.1	96	-0.02
68	Atrazine	0.201	0.179	10.9	93	-0.02
69 C	Pentachlorophenol	0.113	0.111	1.8	92	-0.02
70	Phenanthrene	1.109	1.035	6.7	98	-0.02
71	Anthracene	1.118	1.092	2.3	90	-0.02
72	Carbazole	0.975	0.921	5.5	86	-0.02
73	Di-n-butylphthalate	1.241	1.163	6.3	96	-0.02
74 C	Fluoranthene	1.123	1.087	3.2	91	-0.02
75 I	Chrysene-d12	1.000	1.000	0.0	89	-0.02
76	Benzidine	0.597	0.757	-26.8#	94	-0.02
77	Pyrene	1.531	1.442	5.8	80	-0.03
78 S	Terphenyl-d14	0.883	0.964	-9.2	102	-0.02
79	Butylbenzylphthalate	0.695	0.690	0.7	83	-0.02
80	Benzo(a)anthracene	1.241	1.179	5.0	84	-0.02
81	3,3'-Dichlorobenzidine	0.440	0.450	-2.3	83	-0.03
82	Chrysene	1.204	1.119	7.1	80	-0.03
83	Bis(2-ethylhexyl)phthalate	0.864	0.892	-3.2	89	-0.02
84 c	Di-n-octyl phthalate	1.426	1.390	2.5	84	-0.02
85	Indeno(1,2,3-cd)pyrene	0.947	0.959	-1.3	92	-0.05
86 I	Perylene-d12	1.000	1.000	0.0	88	-0.03
87	Benzo(b)fluoranthene	1.344	1.128	16.1	73	-0.03

Data Path : Z:\HPCHEM1\BNA F\DATA\BF020216\
Data File : BF084760.D
Acq On : 2 Feb 2016 21:55
Operator : SJ/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 03 02:45:28 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF020116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 01 14:55:11 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.111	1.227	-10.4	100	-0.02
89 C	Benzo(a)pyrene	1.145	1.113	2.8	85	-0.02
90	Dibenzo(a,h)anthracene	0.964	0.981	-1.8	91	-0.05
91	Benzo(a,h,i)perylene	0.971	1.023	-5.4	94	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0